

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU031519\
 Data File : VU030064.D
 Acq On : 14 Mar 2019 11:22
 Operator : JC/SP
 Sample : K1784-03
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0962

Quant Time: Mar 15 01:56:26 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM031319WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Mar 15 01:21:27 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	475061	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	454600	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	224783	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	153620	48.89	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	97.78%
7) Chloroethane-d5	1.68	69	122505	48.30	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	96.60%
11) 1,1-Dichloroethene-d2	2.27	63	209572	36.95	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	73.90%
21) 2-Butanone-d5	4.18	46	221632	94.14	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	94.14%
24) Chloroform-d	4.65	84	262343	46.54	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.08%
26) 1,2-Dichloroethane-d4	5.31	65	156058	47.92	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.84%
32) Benzene-d6	5.34	84	567240	48.47	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.94%
36) 1,2-Dichloropropane-d6	6.33	67	175508	49.01	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	98.02%
41) Toluene-d8	7.56	98	520117	47.47	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.94%
43) trans-1,3-Dichloropropene-	7.85	79	81798	47.15	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	94.30%
47) 2-Hexanone-d5	8.31	63	191165	95.39	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	95.39%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	254787	46.30	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	92.60%
64) 1,2-Dichlorobenzene-d4	11.86	152	213852	49.11	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.22%

Target Compounds

					Qvalue
13) Acetone	2.33	43	105572	38.352	ug/L 99
16) Methylene chloride	2.70	84	2350193	668.439	ug/L 99
20) cis-1,2-Dichloroethene	4.23	96	53228	14.419	ug/L 100
22) 2-Butanone	4.28	43	22943	7.321	ug/L 75
25) Chloroform	4.67	83	99357	17.003	ug/L 99
34) Trichloroethene	6.18	95	751771	220.501	ug/L 100
38) Bromodichloromethane	6.76	83	5350	1.257	ug/L 99
40) 4-Methyl-2-pentanone	7.46	43	6942	1.503	ug/L 99
45) 1,1,2-Trichloroethane	8.07	97	5358	1.566	ug/L 90
54) o-xylene	9.77	106	6019	1.017	ug/L 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU031519\
Data File : VU030064.D
Acq On : 14 Mar 2019 11:22
Operator : JC/SP
Sample : K1784-03
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0962

Quant Time: Mar 15 01:56:26 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM031319WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Mar 15 01:21:27 2019
Response via : Initial Calibration

