

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU031519\
 Data File : VU030017.D
 Acq On : 13 Mar 2019 09:47
 Operator : JC/SP
 Sample : VU0315MBL01
 Misc : 5.00g/5mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK58

Quant Time: Mar 14 02:00:27 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM031319WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Mar 14 01:36:14 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	487788	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	472732	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	222609	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	162303	50.30	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	100.60%
7) Chloroethane-d5	1.68	69	134700	51.72	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	103.44%
11) 1,1-Dichloroethene-d2	2.27	63	223372	38.36	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	76.72%
21) 2-Butanone-d5	4.17	46	257234	106.41	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	106.41%
24) Chloroform-d	4.65	84	295404	51.04	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.08%
26) 1,2-Dichloroethane-d4	5.31	65	177140	52.98	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	105.96%
32) Benzene-d6	5.34	84	645771	53.07	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	106.14%
36) 1,2-Dichloropropane-d6	6.33	67	199865	53.67	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	107.34%
41) Toluene-d8	7.56	98	588697	51.66	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	103.32%
43) trans-1,3-Dichloropropene-	7.85	79	92494	51.27	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	102.54%
47) 2-Hexanone-d5	8.31	63	226980	108.92	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	108.92%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	292695	51.14	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	102.28%
64) 1,2-Dichlorobenzene-d4	11.86	152	234497	54.38	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	108.76%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_U\DATA\VU031519\
Data File : VU030017.D
Acq On : 13 Mar 2019 09:47
Operator : JC/SP
Sample : VU0315MBL01
Misc : 5.00g/5mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK58

Quant Time: Mar 14 02:00:27 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM031319WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Mar 14 01:36:14 2019
Response via : Initial Calibration

