

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU031519\
 Data File : VU030052.D
 Acq On : 14 Mar 2019 00:00
 Operator : JC/SP
 Sample : K1885-11
 Misc : 5.54g/5mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BF195

Quant Time: Mar 14 04:54:45 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM031319WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Mar 14 03:50:43 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	507483	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	520442	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.49	152	278529	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	109052	32.49	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	64.98%
7) Chloroethane-d5	1.64	69	54176	19.99	ug/L	-0.04
Spiked Amount	50.000	Range	70 - 130	Recovery	=	39.98%#
11) 1,1-Dichloroethene-d2	2.23	63	158003	26.08	ug/L	-0.04
Spiked Amount	50.000	Range	60 - 125	Recovery	=	52.16%#
21) 2-Butanone-d5	4.19	46	229698	91.33	ug/L	0.02
Spiked Amount	100.000	Range	40 - 130	Recovery	=	91.33%
24) Chloroform-d	4.64	84	235976	39.19	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	78.38%
26) 1,2-Dichloroethane-d4	5.30	65	146257	42.04	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	84.08%
32) Benzene-d6	5.33	84	499776	37.31	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	74.62%
36) 1,2-Dichloropropane-d6	6.32	67	163473	39.87	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	79.74%
41) Toluene-d8	7.56	98	455763	36.33	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	72.66%#
43) trans-1,3-Dichloropropene-	7.85	79	71411	35.95	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	71.90%
47) 2-Hexanone-d5	8.32	63	218557	95.26	ug/L	0.02
Spiked Amount	100.000	Range	45 - 130	Recovery	=	95.26%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	294598	46.76	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	93.52%
64) 1,2-Dichlorobenzene-d4	11.86	152	225140	41.73	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	83.46%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_U\DATA\VU031519\
Data File : VU030052.D
Acq On : 14 Mar 2019 00:00
Operator : JC/SP
Sample : K1885-11
Misc : 5.54g/5mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BF195

Quant Time: Mar 14 04:54:45 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM031319WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Mar 14 03:50:43 2019
Response via : Initial Calibration

