

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU031519\
 Data File : VU030060.D
 Acq On : 14 Mar 2019 09:50
 Operator : JC/SP
 Sample : VU0315WBL03
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK60

Quant Time: Mar 16 00:07:32 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM031319WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Mar 15 01:21:27 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	487539	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	477939	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	230747	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	144982	44.96	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	89.92%
7) Chloroethane-d5	1.68	69	123973	47.62	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	95.24%
11) 1,1-Dichloroethene-d2	2.27	63	209903	36.06	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	72.12%
21) 2-Butanone-d5	4.17	46	257250	106.47	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	106.47%
24) Chloroform-d	4.64	84	290188	50.17	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.34%
26) 1,2-Dichloroethane-d4	5.31	65	175140	52.41	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	104.82%
32) Benzene-d6	5.34	84	634094	51.54	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	103.08%
36) 1,2-Dichloropropane-d6	6.33	67	196787	52.26	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	104.52%
41) Toluene-d8	7.56	98	583790	50.67	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	101.34%
43) trans-1,3-Dichloropropene-	7.85	79	91081	49.94	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	99.88%
47) 2-Hexanone-d5	8.31	63	223812	106.23	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	106.23%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	295612	51.09	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	102.18%
64) 1,2-Dichlorobenzene-d4	11.86	152	242699	54.29	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	108.58%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_U\DATA\VU031519\
Data File : VU030060.D
Acq On : 14 Mar 2019 09:50
Operator : JC/SP
Sample : VU0315WBL03
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK60

Quant Time: Mar 16 00:07:32 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM031319WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Mar 15 01:21:27 2019
Response via : Initial Calibration

