

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031521\
 Data File : VU042653.D
 Acq On : 15 Mar 2021 10:25
 Operator : SY/MD
 Sample : VU0315WBL01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK28

Quant Time: Mar 16 01:32:55 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SOMUTR030221WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Mar 16 01:32:07 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.256	114	83258	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.424	117	81193	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.819	152	36377	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	23145	4.17	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	83.40%
7) Chloroethane-d5	1.919	69	21088	5.02	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	100.40%
11) 1,1-Dichloroethene-d2	2.575	63	41701	3.33	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	66.60%
20) 2-Butanone-d5	4.646	46	87966	48.27	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	96.54%
24) Chloroform-d	5.073	84	54425	4.96	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.20%
26) 1,2-Dichloroethane-d4	5.710	65	36505	5.41	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	108.20%
32) Benzene-d6	5.736	84	95833	4.81	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.20%
36) 1,2-Dichloropropane-d6	6.700	67	32836	5.11	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	102.20%
41) Toluene-d8	7.906	98	90661	4.66	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.20%
43) trans-1,3-Dichloroprop...	8.186	79	13569	4.52	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	90.40%
46) 2-Hexanone-d5	8.645	63	67161	46.75	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	93.50%
57) 1,1,2,2-Tetrachloroeth...	10.764	84	27917	4.88	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	97.60%
64) 1,2-Dichlorobenzene-d4	12.202	152	35720	5.42	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	108.40%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031521\
Data File : VU042653.D
Acq On : 15 Mar 2021 10:25
Operator : SY/MD
Sample : VU0315WBL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK28

Quant Time: Mar 16 01:32:55 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SOMUTR030221WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Mar 16 01:32:07 2021
Response via : Initial Calibration

