

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031521\
 Data File : VU042662.D
 Acq On : 15 Mar 2021 14:10
 Operator : SY/MD
 Sample : M1679-01DL 5X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AF0DL

Quant Time: Mar 16 01:36:21 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SOMUTR030221WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Mar 16 01:32:07 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.256	114	77439	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.424	117	76036	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.819	152	34925	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	20566	3.98	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	79.60%
7) Chloroethane-d5	1.919	69	19055	4.87	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	97.40%
11) 1,1-Dichloroethene-d2	2.572	63	37541	3.22	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	64.40%
20) 2-Butanone-d5	4.649	46	80012	47.20	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	94.40%
24) Chloroform-d	5.073	84	49971	4.89	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.80%
26) 1,2-Dichloroethane-d4	5.713	65	34230	5.45	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	109.00%
32) Benzene-d6	5.736	84	88681	4.76	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.20%
36) 1,2-Dichloropropane-d6	6.700	67	29909	4.97	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	99.40%
41) Toluene-d8	7.906	98	85169	4.68	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.60%
43) trans-1,3-Dichloroprop...	8.189	79	12320	4.39	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	87.80%
46) 2-Hexanone-d5	8.649	63	62059	46.13	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	92.26%
57) 1,1,2,2-Tetrachloroeth...	10.764	84	26114	4.87	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	97.40%
64) 1,2-Dichlorobenzene-d4	12.202	152	33451	5.28	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	105.60%
Target Compounds						Qvalue
47) Tetrachloroethene	8.559	164	27066	5.03	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031521\
Data File : VU042662.D
Acq On : 15 Mar 2021 14:10
Operator : SY/MD
Sample : M1679-01DL 5X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AF0DL

Quant Time: Mar 16 01:36:21 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SOMUTR030221WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Mar 16 01:32:07 2021
Response via : Initial Calibration

