

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031521\
 Data File : VU042665.D
 Acq On : 15 Mar 2021 15:22
 Operator : SY/MD
 Sample : M1676-09DL 5X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AD8DL

Quant Time: Mar 16 01:37:24 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SOMUTR030221WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Mar 16 01:32:07 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.256	114	79862	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.423	117	80151	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.819	152	36369	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	20259	3.81	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	76.20%
7) Chloroethane-d5	1.919	69	18633	4.62	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	92.40%
11) 1,1-Dichloroethene-d2	2.575	63	37890	3.15	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	63.00%
20) 2-Butanone-d5	4.649	46	75846	43.39	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	86.78%
24) Chloroform-d	5.073	84	49194	4.67	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.40%
26) 1,2-Dichloroethane-d4	5.710	65	32339	4.99	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.80%
32) Benzene-d6	5.735	84	87665	4.46	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.20%
36) 1,2-Dichloropropane-d6	6.700	67	29192	4.60	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	92.00%
41) Toluene-d8	7.906	98	83891	4.37	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.40%
43) trans-1,3-Dichloroprop...	8.185	79	11690	3.95	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	79.00%
46) 2-Hexanone-d5	8.645	63	60113	42.39	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	84.78%
57) 1,1,2,2-Tetrachloroeth...	10.764	84	25700	4.55	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	91.00%
64) 1,2-Dichlorobenzene-d4	12.201	152	32850	4.98	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	99.60%
Target Compounds						
47) Tetrachloroethene	8.558	164	41492	7.32	ug/L	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031521\
Data File : VU042665.D
Acq On : 15 Mar 2021 15:22
Operator : SY/MD
Sample : M1676-09DL 5X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AD8DL

Quant Time: Mar 16 01:37:24 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SOMUTR030221WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Mar 16 01:32:07 2021
Response via : Initial Calibration

