

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU031619\
 Data File : VU030113.D
 Acq On : 15 Mar 2019 16:11
 Operator : JC/SP
 Sample : VU0316WBL02
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK33

Quant Time: Mar 16 01:04:08 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM031319WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Mar 16 00:52:41 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	550144	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	517910	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	250997	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	178149	48.96	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	97.92%
7) Chloroethane-d5	1.68	69	146109	49.74	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	99.48%
11) 1,1-Dichloroethene-d2	2.27	63	241675	36.79	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	73.58%
21) 2-Butanone-d5	4.18	46	264759	97.11	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	97.11%
24) Chloroform-d	4.65	84	316861	48.54	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.08%
26) 1,2-Dichloroethane-d4	5.31	65	188596	50.01	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.02%
32) Benzene-d6	5.34	84	681258	51.10	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.20%
36) 1,2-Dichloropropane-d6	6.33	67	213472	52.32	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	104.64%
41) Toluene-d8	7.56	98	609264	48.80	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.60%
43) trans-1,3-Dichloropropene-	7.85	79	95748	48.44	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	96.88%
47) 2-Hexanone-d5	8.31	63	224544	98.35	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	98.35%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	304848	48.62	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	97.24%
64) 1,2-Dichlorobenzene-d4	11.86	152	248025	51.01	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	102.02%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU031619\
Data File : VU030113.D
Acq On : 15 Mar 2019 16:11
Operator : JC/SP
Sample : VU0316WBL02
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK33

Quant Time: Mar 16 01:04:08 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM031319WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Mar 16 00:52:41 2019
Response via : Initial Calibration

