

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU031619\
 Data File : VU030101.D
 Acq On : 15 Mar 2019 11:22
 Operator : JC/SP
 Sample : K1917-19
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C09H9

Quant Time: Mar 16 00:17:14 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM031319WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Mar 16 00:02:21 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	467145	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	451550	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	221216	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	100071	32.39	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	64.78%
7) Chloroethane-d5	1.68	69	94518	37.89	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	75.78%
11) 1,1-Dichloroethene-d2	2.27	63	155092	27.81	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	55.62%#
21) 2-Butanone-d5	4.17	46	236547	102.18	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	102.18%
24) Chloroform-d	4.65	84	250362	45.17	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.34%
26) 1,2-Dichloroethane-d4	5.31	65	151736	47.39	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.78%
32) Benzene-d6	5.34	84	518463	44.60	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.20%
36) 1,2-Dichloropropane-d6	6.33	67	172319	48.44	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	96.88%
41) Toluene-d8	7.56	98	475618	43.70	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	87.40%
43) trans-1,3-Dichloropropene-	7.85	79	77833	45.17	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	90.34%
47) 2-Hexanone-d5	8.31	63	197037	98.99	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	98.99%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	268463	49.11	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	98.22%
64) 1,2-Dichlorobenzene-d4	11.86	152	210032	49.01	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.02%


Target Compounds					Ovalue
13) Acetone	2.32	43	10424	3.851 ug/L	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU031619\
Data File : VU030101.D
Acq On : 15 Mar 2019 11:22
Operator : JC/SP
Sample : K1917-19
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C09H9

Quant Time: Mar 16 00:17:14 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM031319WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Mar 16 00:02:21 2019
Response via : Initial Calibration





#13
 Acetone
 Concen: 3.851 ug/L
 RT: 2.32 min Scan# 383
 Delta R.T. 0.00 min
 Lab File: VU030101.D
 Acq: 15 Mar 2019 11:22

Instrument :
 MSVOA_U
 ClientSampleId :
 C09H9

Tot Ion: 43 Resp: 10424
 Ion Ratio Lower Upper
 43 100
 58 35.3 0.0 73.0

