

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031623\
 Data File : VU053226.D
 Acq On : 16 Mar 2023 15:08
 Operator : JC/MD
 Sample : 01930-05 200X
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 EYFG6

Quant Time: Mar 17 02:42:12 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM031523WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Mar 17 02:34:44 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.243	114	131610	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.414	117	124931	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.809	152	69135	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	58139	47.884	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	95.760%
7) Chloroethane-d5	1.909	69	48851	49.740	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	99.480%
11) 1,1-Dichloroethene-d2	2.562	63	79092	37.603	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	75.200%
21) 2-Butanone-d5	4.613	46	65985	117.173	ug/L	-0.01
Spiked Amount	100.000	Range	40 - 130	Recovery	=	117.170%
24) Chloroform-d	5.057	84	101724	50.867	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.740%
26) 1,2-Dichloroethane-d4	5.697	65	66926	53.887	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	107.780%
32) Benzene-d6	5.722	84	213806	52.763	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	105.520%
36) 1,2-Dichloropropane-d6	6.684	67	70908	54.218	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	108.440%
41) Toluene-d8	7.893	98	190419	50.473	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.940%
43) trans-1,3-Dichloroprop...	8.176	79	32634	50.575	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	101.140%
47) 2-Hexanone-d5	8.626	63	49131	120.162	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	120.160%
56) 1,1,2,2-Tetrachloroeth...	10.751	84	92232	54.440	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	108.880%
66) 1,2-Dichlorobenzene-d4	12.188	152	61409	45.874	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.740%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031623\
Data File : VU053226.D
Acq On : 16 Mar 2023 15:08
Operator : JC/MD
Sample : 01930-05 200X
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
EYFG6

Quant Time: Mar 17 02:42:12 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM031523WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Mar 17 02:34:44 2023
Response via : Initial Calibration

