

Data File : VU037321.D
Acq On : 17 Mar 2020 16:52
Operator : JC/MD
Sample : L1494-02
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :

Quant Time: Mar 18 04:22:09 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM031220WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Mar 18 02:36:04 2020
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.28	114	416001	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.44	117	402027	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.83	152	161698	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	122398	38.02	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	76.04%
7) Chloroethane-d5	1.93	69	114408	44.08	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	88.16%
11) 1,1-Dichloroethene-d2	2.59	63	187894	31.81	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	63.62%
21) 2-Butanone-d5	4.67	46	267068	104.82	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	104.82%
24) Chloroform-d	5.11	84	286654	46.31	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.62%
26) 1,2-Dichloroethane-d4	5.74	65	212924	46.99	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.98%
32) Benzene-d6	5.76	84	525312	47.47	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.94%
36) 1,2-Dichloropropane-d6	6.72	67	175215	48.37	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	96.74%
41) Toluene-d8	7.93	98	494370	46.73	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.46%
43) trans-1,3-Dichloropropene-	8.20	79	90909	46.10	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	92.20%
47) 2-Hexanone-d5	8.66	63	188456	100.48	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	100.48%
57) 1,1,2,2-Tetrachloroethane-	10.77	84	267734	47.71	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	95.42%
64) 1,2-Dichlorobenzene-d4	12.21	152	182484	54.41	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	108.82%

Target Compounds					Qvalue
12) 1,1-Dichloroethene	2.60	96	7724	3.146 ug/L	# 1
13) Acetone	2.65	43	4795m	1.984 ug/L	
20) cis-1,2-Dichloroethene	4.71	96	8908	2.863 ug/L	74
30) 1,1,1-Trichloroethane	5.35	97	6885	1.290 ug/L	97
34) Trichloroethene	6.57	95	113571	36.883 ug/L	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VU037321.D
Acq On : 17 Mar 2020 16:52
Operator : JC/MD
Sample : L1494-02
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :

Quant Time: Mar 18 04:22:09 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM031220WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Mar 18 02:36:04 2020
Response via : Initial Calibration

