

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU031720\
 Data File : VU037324.D
 Acq On : 17 Mar 2020 18:00
 Operator : JC/MD
 Sample : L1494-05
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 CONH6

Quant Time: Mar 18 04:30:06 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM031220WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Mar 18 02:36:04 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.28	114	427691	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.44	117	424497	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.83	152	169143	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	122743	37.08	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	74.16%
7) Chloroethane-d5	1.93	69	112094	42.00	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	84.00%
11) 1,1-Dichloroethene-d2	2.59	63	176242	29.02	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	58.04%#
21) 2-Butanone-d5	4.67	46	263004	100.40	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	100.40%
24) Chloroform-d	5.11	84	283092	44.48	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.96%
26) 1,2-Dichloroethane-d4	5.74	65	207233	44.49	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.98%
32) Benzene-d6	5.76	84	517334	44.27	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.54%
36) 1,2-Dichloropropane-d6	6.72	67	172753	45.16	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	90.32%
41) Toluene-d8	7.93	98	488464	43.73	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	87.46%
43) trans-1,3-Dichloropropene-	8.20	79	88090	42.31	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	84.62%
47) 2-Hexanone-d5	8.66	63	180486	91.13	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	91.13%
57) 1,1,2,2-Tetrachloroethane-	10.77	84	265500	44.81	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	89.62%
64) 1,2-Dichlorobenzene-d4	12.21	152	178893	50.99	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	101.98%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU031720\
Data File : VU037324.D
Acq On : 17 Mar 2020 18:00
Operator : JC/MD
Sample : L1494-05
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
CONH6

Quant Time: Mar 18 04:30:06 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM031220WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Mar 18 02:36:04 2020
Response via : Initial Calibration

