


Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031721\
Data File : VU042684.D
Acq On : 17 Mar 2021 09:36
Operator : SY/MD
Sample : BFB
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BFB

Integration File: rteint.p

Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U031721DW.M
Title : METHOD 524.2 VOLATILES DRINKING WATER
Last Update : Thu Mar 18 07:25:09 2021



AutoFind: Averaged scan 2880 to 2918; Bkg corrected with scan 2879

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	24.1	3286	PASS
75	95	30	60	58.3	7943	PASS
95	95	100	100	100.0	13625	PASS
96	95	5	9	6.7	911	PASS
173	174	0.00	2	0.1	7	PASS
174	95	50	100	92.0	12535	PASS
175	174	5	9	7.8	981	PASS
176	174	95	101	96.1	12041	PASS
177	176	5	9	6.7	809	PASS