

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031723\
 Data File : VU053284.D
 Acq On : 17 Mar 2023 22:22
 Operator : JC/MD
 Sample : 01961-01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 EYFB6

Quant Time: Mar 18 04:58:59 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM031523WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Mar 18 04:29:10 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.243	114	137883	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.414	117	128357	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.806	152	57405	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	52888	41.577	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	83.160%	
7) Chloroethane-d5	1.912	69	44945	43.681	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	87.360%	
11) 1,1-Dichloroethene-d2	2.565	63	68529	31.099	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	62.200%	
21) 2-Butanone-d5	4.610	46	65587	111.168	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	111.170%	
24) Chloroform-d	5.057	84	85941	41.019	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	82.040%	
26) 1,2-Dichloroethane-d4	5.697	65	66404	51.034	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	102.060%	
32) Benzene-d6	5.723	84	197824	47.516	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	95.040%	
36) 1,2-Dichloropropane-d6	6.684	67	67944	50.565	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	101.140%	
41) Toluene-d8	7.893	98	179235	46.240	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	92.480%	
43) trans-1,3-Dichloroprop...	8.176	79	29308	44.208	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	88.420%	
47) 2-Hexanone-d5	8.626	63	48050	114.381	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	114.380%	
56) 1,1,2,2-Tetrachloroeth...	10.751	84	86403	49.639	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	99.280%	
66) 1,2-Dichlorobenzene-d4	12.188	152	57164	51.429	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	102.860%	
Target Compounds						
25) Chloroform	5.083	83	9052	4.313	ug/L	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031723\
Data File : VU053284.D
Acq On : 17 Mar 2023 22:22
Operator : JC/MD
Sample : 01961-01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
EYFB6

Quant Time: Mar 18 04:58:59 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM031523WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Mar 18 04:29:10 2023
Response via : Initial Calibration

