

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031723\
 Data File : VU053266.D
 Acq On : 17 Mar 2023 14:50
 Operator : JC/MD
 Sample : 01930-07
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 EYFB2

Quant Time: Mar 18 01:52:20 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM031523WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Mar 18 01:49:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.243	114	124624	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.414	117	113860	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.809	152	53342	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	44052	38.315	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	76.640%
7) Chloroethane-d5	1.912	69	37524	40.349	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	80.700%
11) 1,1-Dichloroethene-d2	2.565	63	60527	30.390	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	60.780%
21) 2-Butanone-d5	4.613	46	58519	109.741	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	109.740%
24) Chloroform-d	5.057	84	84113	44.418	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.840%
26) 1,2-Dichloroethane-d4	5.697	65	56566	48.098	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.200%
32) Benzene-d6	5.723	84	174978	47.380	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.760%
36) 1,2-Dichloropropane-d6	6.684	67	60074	50.401	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	100.800%
41) Toluene-d8	7.893	98	156241	45.440	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	90.880%
43) trans-1,3-Dichloroprop...	8.176	79	26674	45.358	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	90.720%
47) 2-Hexanone-d5	8.626	63	42090	112.951	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	112.950%
56) 1,1,2,2-Tetrachloroeth...	10.751	84	78127	50.599	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	101.200%
66) 1,2-Dichlorobenzene-d4	12.189	152	50889	49.270	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.540%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031723\
Data File : VU053266.D
Acq On : 17 Mar 2023 14:50
Operator : JC/MD
Sample : 01930-07
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
EYFB2

Quant Time: Mar 18 01:52:20 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM031523WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Mar 18 01:49:30 2023
Response via : Initial Calibration

