

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031723\
 Data File : VU053274.D
 Acq On : 17 Mar 2023 18:00
 Operator : JC/MD
 Sample : 01930-18
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 EYFF6

Quant Time: Mar 18 01:54:05 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM031523WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Mar 18 01:49:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.243	114	141657	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.414	117	130189	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.809	152	59325	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	56062	42.898	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	85.800%	
7) Chloroethane-d5	1.912	69	47842	45.258	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	90.520%	
11) 1,1-Dichloroethene-d2	2.565	63	72024	31.814	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	63.620%	
21) 2-Butanone-d5	4.613	46	65570	108.178	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	108.180%	
24) Chloroform-d	5.057	84	96445	44.806	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	89.620%	
26) 1,2-Dichloroethane-d4	5.697	65	66920	50.061	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	100.120%	
32) Benzene-d6	5.722	84	206623	48.931	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	97.860%	
36) 1,2-Dichloropropane-d6	6.684	67	70736	51.902	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	103.800%	
41) Toluene-d8	7.893	98	186149	47.348	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	94.700%	
43) trans-1,3-Dichloroprop...	8.176	79	31376	46.661	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	93.320%	
47) 2-Hexanone-d5	8.626	63	50161	117.726	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	117.730%	
56) 1,1,2,2-Tetrachloroeth...	10.751	84	90742	51.398	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	102.800%	
66) 1,2-Dichlorobenzene-d4	12.188	152	58453	50.886	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	101.780%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031723\
Data File : VU053274.D
Acq On : 17 Mar 2023 18:00
Operator : JC/MD
Sample : 01930-18
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
EYFF6

Quant Time: Mar 18 01:54:05 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM031523WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Mar 18 01:49:30 2023
Response via : Initial Calibration

