

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031821\
 Data File : VU042700.D
 Acq On : 18 Mar 2021 10:11
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_U
 LabSampleId :
 VSTDCCC010

Quant Time: Mar 19 02:17:31 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U031721DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Thu Mar 18 07:25:09 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 200%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 i	Fluorobenzene	1.000	1.000	0.0	100	0.00
2 T	Dichlorodifluoromethane	0.389	0.395	-1.5	103	0.00
3 t	Chloromethane	0.368	0.358	2.7	104	0.00
4 Rt	Vinyl Chloride	0.355	0.354	0.3	102	0.00
5 T	Bromomethane	0.234	0.247	-5.6	102	0.00
6 T	Chloroethane	0.242	0.244	-0.8	102	0.00
7 T	Trichlorofluoromethane	0.606	0.621	-2.5	103	0.00
8	1,1,2-Trichloro-1,2,2-trifl	0.304	0.314	-3.3	105	0.00
9 Rt	1,1-Dichloroethene	0.263	0.267	-1.5	105	0.00
10 t	Iodomethane	0.414	0.431	-4.1	103	0.00
11 t	Allyl Chloride	0.503	0.508	-1.0	106	0.00
12 t	Acrylonitrile	0.076	0.081	-6.6	106	0.00
13 T	Acetone	0.059	0.057	3.4	98	0.00
14 T	Carbon Disulfide	0.824	0.827	-0.4	103	0.00
15 RT	Methylene Chloride	0.319	0.305	4.4	104	0.00
16 RT	trans-1,2-Dichloroethene	0.284	0.291	-2.5	107	0.00
17 t	1,1-Dichloroethane	0.583	0.590	-1.2	105	0.00
18 T	2-Butanone	0.109	0.125	-14.7	118	0.00
19	Cyclohexane	0.497	0.490	1.4	102	0.00
20	Methylcyclohexane	0.431	0.491	-13.9	109	0.00
21 T	2,2-Dichloropropane	0.536	0.547	-2.1	106	0.00
22 RT	cis-1,2-Dichloroethene	0.320	0.323	-0.9	106	0.00
23 t	Diethyl Ether	0.223	0.226	-1.3	102	0.00
24 t	tert-Butyl Alcohol	0.016	0.018	-12.5	97	0.00
25 t	Methyl tert-Butyl Ether	0.779	0.785	-0.8	105	0.00
26 t	Bromochloromethane	0.153	0.148	3.3	98	0.00
27 t	Chloroform	0.608	0.570	6.3	97	0.00
28 RT	1,1,1-Trichloroethane	0.552	0.516	6.5	96	0.00
29 T	1,1-Dichloropropene	0.387	0.416	-7.5	108	0.00
30 RT	Carbon Tetrachloride	0.428	0.467	-9.1	108	0.00
31 t	Isopropyl Ether	1.042	1.040	0.2	104	0.00
32	Ethyl-t-butyl ether	0.927	0.953	-2.8	106	0.00
33	Tert-Amyl methyl ether	0.674	0.729	-8.2	105	0.00
34 t	Propionitrile	0.024	0.027	-12.5	97	0.00
35 RT	Benzene	1.087	1.157	-6.4	108	0.00
36 RT	1,2-Dichloroethane	0.421	0.445	-5.7	110	0.00
37 RT	Trichloroethene	0.292	0.316	-8.2	108	0.00
38 Rt	1,2-Dichloropropane	0.312	0.328	-5.1	107	0.00
39 t	Methacrylonitrile	0.151	0.131	13.2	90	0.00
40 t	Methyl acrylate	0.197	0.219	-11.2	103	0.00
41 t	Tetrahydrofuran	0.075	0.070	6.7	99	0.00
42 t	1-Chlorobutane	0.572	0.602	-5.2	109	0.00
43 T	Dibromomethane	0.171	0.184	-7.6	104	0.00
44 T	Bromodichloromethane	0.401	0.438	-9.2	108	0.00
45 T	4-Methyl-2-Pentanone	0.257	0.282	-9.7	104	0.00
46 t	t-1,4-Dichloro-2-butene	0.083	0.106	-27.7	104	0.00
47 t	Methyl methacrylate	0.150	0.168	-12.0	105	0.00
48 t	Ethyl methacrylate	0.298	0.331	-11.1	103	0.00
49 Rt	Toluene	0.674	0.768	-13.9	107	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031821\
 Data File : VU042700.D
 Acq On : 18 Mar 2021 10:11
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_U
 LabSampleId :
 VSTDCCC010

Quant Time: Mar 19 02:17:31 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U031721DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Thu Mar 18 07:25:09 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 200%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
50 T	t-1,3-Dichloropropene	0.360	0.420	-16.7	108	0.00
51 T	cis-1,3-Dichloropropene	0.402	0.447	-11.2	107	0.00
52 RT	1,1,2-Trichloroethane	0.236	0.242	-2.5	103	0.00
53 t	1,3-Dichloropropane	0.407	0.433	-6.4	105	0.00
54 t	2-Hexanone	0.183	0.206	-12.6	103	0.00
55 t	Dibromochloromethane	0.281	0.308	-9.6	106	0.00
56 T	1,2-Dibromoethane	0.235	0.255	-8.5	108	0.00
57 S	4-Bromofluorobenzene	0.445	0.532	-19.6	109	0.00
58 RT	Tetrachloroethene	0.317	0.337	-6.3	107	0.00
59 Rt	Chlorobenzene	0.761	0.834	-9.6	106	0.00
60 T	1,1,1,2-Tetrachloroethane	0.298	0.328	-10.1	104	0.00
61 t	Pentachloroethane	0.229	0.252	-10.0	104	0.00
62 t	Hexachloroethane	0.237	0.281	-18.6	107	0.00
63 Rt	Ethyl Benzene	1.293	1.514	-17.1	106	0.00
64 RT	m/p-Xylenes	0.498	0.596	-19.7	107	0.00
65 RT	o-Xylene	0.482	0.568	-17.8	107	0.00
66 RT	Styrene	0.825	1.001	-21.3	106	0.00
67 t	Bromoform	0.183	0.205	-12.0	103	0.00
68 S	1,2-Dichlorobenzene-d4	0.471	0.513	-8.9	102	0.00
69 T	Isopropylbenzene	1.311	1.566	-19.5	108	0.00
70 T	1,1,2,2-Tetrachloroethane	0.312	0.334	-7.1	102	0.00
71 T	1,2,3-Trichloropropane	0.246	0.242	1.6	92	0.00
72 t	Bromobenzene	0.368	0.415	-12.8	108	0.00
73 t	n-propylbenzene	0.368	0.437	-18.8	105	0.00
74 t	2-Chlorotoluene	0.334	0.386	-15.6	106	0.00
75 t	1,3,5-Trimethylbenzene	1.174	1.433	-22.1	106	0.00
76 t	4-Chlorotoluene	0.354	0.417	-17.8	105	0.00
77 t	tert-Butylbenzene	1.167	1.379	-18.2	106	0.00
78 t	1,2,4-Trimethylbenzene	1.212	1.478	-21.9	106	0.00
79 t	sec-Butylbenzene	1.573	1.874	-19.1	106	0.00
80	Nitrobenzene	0.008	0.009	-12.5	96	0.00
81 t	p-Isopropyltoluene	1.324	1.628	-23.0	107	0.00
82 t	1,3-Dichlorobenzene	0.735	0.823	-12.0	104	0.00
83 Rt	1,4-Dichlorobenzene	0.739	0.847	-14.6	106	0.00
84 t	n-Butylbenzene	1.270	1.590	-25.2	108	0.00
85 Rt	1,2-Dichlorobenzene	0.722	0.828	-14.7	107	0.00
86 t	1,2-Dibromo-3-Chloropropane	0.058	0.069	-19.0	103	0.00
87 Rt	1,2,4-Trichlorobenzene	0.484	0.568	-17.4	105	0.00
88 t	Hexachlorobutadiene	0.325	0.361	-11.1	104	0.00
89 t	Naphthalene	0.791	0.946	-19.6	104	0.00
90 t	1,2,3-Trichlorobenzene	0.460	0.548	-19.1	106	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0