


Iodomethane

Response Ratio



$$\text{Response} = 1.595\text{e-}001 * \text{Amt} - 1.803\text{e-}001$$


Coef of Det (r^2) = 0.994615 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA U\Method\524U031822DW.M

Calibration Table Last Updated: Fri Mar 18 13:19:37 2022

n-Butylbenzene

Response Ratio



$$\text{Response} = 1.359\text{e}+000 * \text{Amt} - 6.162\text{e}-001$$


Coef of Det (r^2) = 0.997826 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA U\Method\524U031822DW.M

Calibration Table Last Updated: Fri Mar 18 13:19:37 2022

p-Isopropyltoluene

Response Ratio



$$\text{Response} = 1.444\text{e}+000 * \text{Amt} - 6.181\text{e}-001$$


Coef of Det (r^2) = 0.998653 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA U\Method\524U031822DW.M

Calibration Table Last Updated: Fri Mar 18 13:19:37 2022

1,2,4-Trimethylbenzene

Response Ratio



$$\text{Response} = 1.276\text{e}+000 * \text{Amt} - 4.424\text{e}-001$$


Coef of Det (r^2) = 0.999167 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA U\Method\524U031822DW.M

Calibration Table Last Updated: Fri Mar 18 13:19:37 2022

1,3,5-Trimethylbenzene

Response Ratio



$$\text{Response} = 1.248\text{e}+000 * \text{Amt} - 4.355\text{e}-001$$


Coef of Det (r^2) = 0.999218 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA U\Method\524U031822DW.M

Calibration Table Last Updated: Fri Mar 18 13:19:37 2022

Propionitrile

Response Ratio



$$\text{Response} = 1.807\text{e-}002 * \text{Amt} + 6.053\text{e-}003$$


Coef of Det (r^2) = 0.994566 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA U\Method\524U031822DW.M

Calibration Table Last Updated: Fri Mar 18 13:19:37 2022

Methylene Chloride

Response Ratio



$$\text{Response} = 2.891\text{e-}001 * \text{Amt} + 1.774\text{e-}001$$

Coef of Det (r^2) = 0.999318 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA U\Method\524U031822DW.M

Calibration Table Last Updated: Fri Mar 18 13:19:37 2022