

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031822\
 Data File : VU047622.D
 Acq On : 18 Mar 2022 18:40
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_U
 LabSampleId :
 VSTDCCC010

Quant Time: Mar 19 05:47:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U031822DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Fri Mar 18 13:19:37 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 200%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 i	Fluorobenzene	1.000	1.000	0.0	95	0.00
2 T	Dichlorodifluoromethane	0.361	0.346	4.2	94	0.00
3 t	Chloromethane	0.353	0.339	4.0	95	0.00
4 Rt	Vinyl Chloride	0.343	0.332	3.2	95	0.00
5 T	Bromomethane	0.201	0.188	6.5	93	0.00
6 T	Chloroethane	0.205	0.198	3.4	95	0.00
7 T	Trichlorofluoromethane	0.474	0.463	2.3	96	0.00
8	1,1,2-Trichloro-1,2,2-trifl	0.271	0.265	2.2	96	0.00
9 Rt	1,1-Dichloroethene	0.267	0.259	3.0	96	0.00
10 t	Iodomethane	0.109	0.161	-47.7#	116	0.00
11 t	Allyl Chloride	0.336	0.323	3.9	95	0.00
12 t	Acrylonitrile	0.059	0.061	-3.4	96	0.00
13 T	Acetone	0.040	0.037	7.5	95	0.00
14 T	Carbon Disulfide	0.845	0.818	3.2	95	0.00
15 RT	Methylene Chloride	0.389	0.315	19.0	100	0.00
16 RT	trans-1,2-Dichloroethene	0.281	0.284	-1.1	98	0.00
17 t	1,1-Dichloroethane	0.477	0.472	1.0	96	0.00
18 T	2-Butanone	0.063	0.065	-3.2	95	0.00
19	Cyclohexane	0.358	0.373	-4.2	93	0.00
20	Methylcyclohexane	0.411	0.433	-5.4	94	0.00
21 T	2,2-Dichloropropane	0.420	0.407	3.1	95	0.00
22 RT	cis-1,2-Dichloroethene	0.305	0.306	-0.3	96	0.00
23 t	Diethyl Ether	0.195	0.189	3.1	95	0.00
24 t	tert-Butyl Alcohol	0.013	0.013	0.0	99	-0.01
25 t	Methyl tert-Butyl Ether	0.651	0.665	-2.2	98	0.00
26 t	Bromochloromethane	0.154	0.150	2.6	98	0.00
27 t	Chloroform	0.516	0.519	-0.6	98	0.00
28 RT	1,1,1-Trichloroethane	0.448	0.449	-0.2	98	0.00
29 T	1,1-Dichloropropene	0.363	0.373	-2.8	95	0.00
30 RT	Carbon Tetrachloride	0.399	0.405	-1.5	96	0.00
31 t	Isopropyl Ether	0.634	0.654	-3.2	96	0.00
32	Ethyl-t-butyl ether	0.640	0.671	-4.8	97	0.00
33	Tert-Amyl methyl ether	0.601	0.632	-5.2	95	0.00
34 t	Propionitrile	0.017	0.018	-5.9	99	0.00
35 RT	Benzene	1.086	1.105	-1.7	97	0.00
36 RT	1,2-Dichloroethane	0.329	0.329	0.0	96	0.00
37 RT	Trichloroethene	0.314	0.313	0.3	95	0.00
38 Rt	1,2-Dichloropropane	0.280	0.281	-0.4	96	0.00
39 t	Methacrylonitrile	0.078	0.080	-2.6	95	0.00
40 t	Methyl acrylate	0.132	0.149	-12.9	100	0.00
41 t	Tetrahydrofuran	0.040	0.039	2.5	92	0.00
42 t	1-Chlorobutane	0.468	0.481	-2.8	97	0.00
43 T	Dibromomethane	0.168	0.165	1.8	96	0.00
44 T	Bromodichloromethane	0.389	0.388	0.3	98	0.00
45 T	4-Methyl-2-Pentanone	0.168	0.171	-1.8	93	0.00
46 t	t-1,4-Dichloro-2-butene	0.073	0.084	-15.1	105	0.00
47 t	Methyl methacrylate	0.136	0.145	-6.6	92	0.00
48 t	Ethyl methacrylate	0.256	0.284	-10.9	97	0.00
49 Rt	Toluene	0.668	0.709	-6.1	96	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031822\
 Data File : VU047622.D
 Acq On : 18 Mar 2022 18:40
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_U
 LabSampleId :
 VSTDCCC010

Quant Time: Mar 19 05:47:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U031822DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Fri Mar 18 13:19:37 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 200%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
50 T	t-1,3-Dichloropropene	0.360	0.372	-3.3	97	0.00
51 T	cis-1,3-Dichloropropene	0.397	0.400	-0.8	94	0.00
52 RT	1,1,2-Trichloroethane	0.233	0.240	-3.0	99	0.00
53 t	1,3-Dichloropropane	0.380	0.389	-2.4	99	0.00
54 t	2-Hexanone	0.122	0.132	-8.2	99	0.00
55 t	Dibromochloromethane	0.295	0.297	-0.7	97	0.00
56 T	1,2-Dibromoethane	0.232	0.234	-0.9	98	0.00
57 S	4-Bromofluorobenzene	0.373	0.369	1.1	96	0.00
58 RT	Tetrachloroethene	0.289	0.285	1.4	94	0.00
59 Rt	Chlorobenzene	0.778	0.800	-2.8	96	0.00
60 T	1,1,1,2-Tetrachloroethane	0.302	0.302	0.0	96	0.00
61 t	Pentachloroethane	0.256	0.256	0.0	97	0.00
62 t	Hexachloroethane	0.254	0.258	-1.6	98	0.00
63 Rt	Ethyl Benzene	1.219	1.321	-8.4	95	0.00
64 RT	m/p-Xylenes	0.484	0.531	-9.7	96	0.00
65 RT	o-Xylene	0.458	0.503	-9.8	96	0.00
66 RT	Styrene	0.789	0.896	-13.6	97	0.00
67 t	Bromoform	0.196	0.204	-4.1	99	0.00
68 S	1,2-Dichlorobenzene-d4	0.452	0.469	-3.8	96	0.00
69 T	Isopropylbenzene	1.212	1.333	-10.0	96	0.00
70 T	1,1,2,2-Tetrachloroethane	0.311	0.315	-1.3	98	0.00
71 T	1,2,3-Trichloropropane	0.245	0.247	-0.8	94	0.00
72 t	Bromobenzene	0.363	0.378	-4.1	95	0.00
73 t	n-propylbenzene	0.352	0.394	-11.9	96	0.00
74 t	2-Chlorotoluene	0.323	0.353	-9.3	98	0.00
75 t	1,3,5-Trimethylbenzene	1.032	1.167	-13.1	96	0.00
76 t	4-Chlorotoluene	0.344	0.379	-10.2	96	0.00
77 t	tert-Butylbenzene	1.055	1.157	-9.7	96	0.00
78 t	1,2,4-Trimethylbenzene	1.055	1.192	-13.0	96	0.00
79 t	sec-Butylbenzene	1.376	1.527	-11.0	95	0.00
80	Nitrobenzene	0.013	0.013	0.0	94	0.00
81 t	p-Isopropyltoluene	1.159	1.309	-12.9	95	0.00
82 t	1,3-Dichlorobenzene	0.718	0.738	-2.8	96	0.00
83 Rt	1,4-Dichlorobenzene	0.722	0.759	-5.1	98	0.00
84 t	n-Butylbenzene	1.089	1.205	-10.7	95	0.00
85 Rt	1,2-Dichlorobenzene	0.689	0.726	-5.4	99	0.00
86 t	1,2-Dibromo-3-Chloropropane	0.052	0.054	-3.8	98	0.00
87 Rt	1,2,4-Trichlorobenzene	0.492	0.535	-8.7	96	0.00
88 t	Hexachlorobutadiene	0.296	0.302	-2.0	95	0.00
89 t	Naphthalene	0.836	0.916	-9.6	94	0.00
90 t	1,2,3-Trichlorobenzene	0.464	0.511	-10.1	96	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0