

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU031919\
 Data File : VU030196.D
 Acq On : 19 Mar 2019 16:00
 Operator : JC/SP
 Sample : K1918-21DL 40X
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C09A5DL

Quant Time: Mar 20 03:58:35 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM031919WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Mar 20 02:58:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	542811	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	525154	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	257157	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	193242	52.65	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	105.30%
7) Chloroethane-d5	1.68	69	157757	53.75	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	107.50%
11) 1,1-Dichloroethene-d2	2.27	63	261679	39.08	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	78.16%
21) 2-Butanone-d5	4.17	46	271735	99.79	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	99.79%
24) Chloroform-d	4.64	84	334923	49.72	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.44%
26) 1,2-Dichloroethane-d4	5.31	65	197027	51.28	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.56%
32) Benzene-d6	5.34	84	732024	52.90	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	105.80%
36) 1,2-Dichloropropane-d6	6.33	67	223543	52.21	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	104.42%
41) Toluene-d8	7.56	98	664177	51.04	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	102.08%
43) trans-1,3-Dichloropropene-	7.85	79	102426	49.12	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	98.24%
47) 2-Hexanone-d5	8.31	63	226473	101.99	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	101.99%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	324132	49.35	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	98.70%
64) 1,2-Dichlorobenzene-d4	11.85	152	265371	51.40	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	102.80%

Target Compounds

					Ovalue
51) Chlorobenzene	9.12	112	74484	6.827 ug/L	98
62) 1,3-Dichlorobenzene	11.41	146	101241	12.340 ug/L	98
63) 1,4-Dichlorobenzene	11.50	146	238805	28.061 ug/L	99
65) 1,2-Dichlorobenzene	11.88	146	97228	11.599 ug/L	99
68) 1,2,4-trichlorobenzene	13.50	180	231473	40.822 ug/L	99
70) 1,2,3-Trichlorobenzene	13.99	180	22436	3.792 ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU031919\
Data File : VU030196.D
Acq On : 19 Mar 2019 16:00
Operator : JC/SP
Sample : K1918-21DL 40X
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampled :
C09A5DL

Quant Time: Mar 20 03:58:35 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM031919WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Mar 20 02:58:59 2019
Response via : Initial Calibration

