

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU031919\
 Data File : VU030192.D
 Acq On : 19 Mar 2019 14:27
 Operator : JC/SP
 Sample : VU0319MBL01
 Misc : 5.00g/5mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK39

Quant Time: Mar 20 03:43:23 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM031919WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Mar 20 02:58:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	547348	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	530822	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	253192	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	175219	47.34	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	94.68%
7) Chloroethane-d5	1.68	69	149932	50.66	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	101.32%
11) 1,1-Dichloroethene-d2	2.27	63	243799	36.11	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	72.22%
21) 2-Butanone-d5	4.17	46	282829	103.01	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	103.01%
24) Chloroform-d	4.65	84	325320	47.90	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.80%
26) 1,2-Dichloroethane-d4	5.31	65	196246	50.66	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.32%
32) Benzene-d6	5.34	84	712261	50.92	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.84%
36) 1,2-Dichloropropane-d6	6.33	67	221113	51.09	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	102.18%
41) Toluene-d8	7.56	98	646843	49.18	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.36%
43) trans-1,3-Dichloropropene-	7.85	79	105174	49.90	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	99.80%
47) 2-Hexanone-d5	8.31	63	234700	104.57	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	104.57%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	328152	49.43	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	98.86%
64) 1,2-Dichlorobenzene-d4	11.86	152	261468	51.44	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	102.88%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU031919\
Data File : VU030192.D
Acq On : 19 Mar 2019 14:27
Operator : JC/SP
Sample : VU0319MBL01
Misc : 5.00µg/5mL/100µL/5.0mL/MSVOA_U/MEOH
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK39

Quant Time: Mar 20 03:43:23 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM031919WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Mar 20 02:58:59 2019
Response via : Initial Calibration

