


Data Path : Z:\voasrv\HPCHEM1\MSVOA U\Data\VU031919\
 Data File : VU030182.D
 Acq On : 19 Mar 2019 10:20
 Operator : JC/SP
 Sample : BFB37
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFB37

Integration File: LSCINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM031919WMA.M
 Title : VOC Analysis
 Last Update : Wed Mar 20 02:58:59 2019



AutoFind: Scans 2865, 2866, 2867; Background Corrected with Scan 2853

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	15.9	12217	PASS
75	95	30	80	45.5	34834	PASS
95	95	100	100	100.0	76624	PASS
96	95	5	9	6.5	4948	PASS
173	174	0.00	2	0.8	566	PASS
174	95	50	120	91.5	70090	PASS
175	174	5	9	6.8	4747	PASS
176	174	95	101	95.2	66749	PASS
177	176	5	9	6.5	4346	PASS

Scan 2853 (10.263 min): VU030182.D
BFB37

m/z	Abundance
40.00	7249.0
43.90	443.0

Instrument :
MSVOA_U
ClientSampleId :
BFB37

m/z	Abundance
35.90	537.0
37.00	2640.0
38.00	2474.0
39.00	853.0
40.00	7053.0
44.00	709.0
44.90	521.0
46.20	101.0
47.00	578.0
48.00	336.0
49.10	2478.0
50.00	12044.0
51.10	3355.0
52.00	132.0
54.90	159.0
56.00	1030.0
57.10	1376.0
60.00	482.0
61.00	2946.0
62.00	3170.0
63.00	2526.0
64.10	246.0
65.00	688.0
67.00	239.0
68.00	6617.0
69.00	6782.0
69.90	573.0
72.00	451.0
73.00	3033.0
74.00	11035.0
75.00	34368.0
76.00	2925.0
77.00	429.0
78.00	296.0
78.90	1864.0
80.00	549.0
80.90	1849.0
81.90	600.0
85.90	118.0
87.00	3229.0
88.00	3441.0
91.00	166.0
92.10	1902.0
93.00	2860.0
94.00	8208.0
95.10	75056.0
96.00	4838.0
103.90	310.0
105.80	307.0
115.90	214.0
116.90	455.0
117.90	231.0
118.80	255.0
127.90	311.0
129.90	231.0
130.90	118.0
134.90	124.0
140.90	531.0
142.90	890.0
144.90	215.0
146.00	185.0
147.90	144.0
154.90	240.0
172.00	959.0
172.90	554.0
174.00	66536.0
175.00	4951.0
176.00	63184.0
177.00	4205.0

Instrument :
MSVOA_U
ClientSampleId :
BFB37

m/z	Abundance
35.90	525.0
37.00	2860.0
38.00	2522.0
39.00	900.0
40.00	7377.0
44.00	717.0
44.90	561.0
46.20	146.0
47.10	670.0
48.00	369.0
49.10	2459.0
50.00	12721.0
51.00	3573.0
52.00	124.0
54.90	132.0
56.00	1004.0
57.00	1468.0
60.00	538.0
61.10	3054.0
62.00	3161.0
63.00	2737.0
64.00	323.0
65.00	777.0
66.90	236.0
68.00	6915.0
69.00	7172.0
69.90	565.0
72.00	530.0
73.00	3107.0
74.00	11772.0
75.00	35912.0
76.00	3006.0
77.00	554.0
77.80	338.0
79.00	1988.0
80.00	695.0
80.90	2041.0
81.90	617.0
85.90	170.0
87.00	3623.0
88.00	3392.0
91.00	208.0
92.10	1938.0
93.00	2901.0
94.00	8943.0
95.00	78776.0
96.00	5135.0
97.10	112.0
103.90	408.0
105.90	274.0
116.00	271.0
116.90	485.0
117.90	300.0
118.90	351.0
127.90	313.0
129.00	113.0
129.90	273.0
130.90	147.0
134.80	133.0
140.90	574.0
142.90	840.0
144.90	236.0
145.90	131.0
148.00	169.0
149.70	119.0
154.90	250.0
172.00	999.0
172.90	573.0
174.00	72600.0
175.00	4831.0
176.00	68920.0
177.00	4564.0
207.00	100.0

Instrument :
MSVOA_U
ClientSampleId :
BFB37

m/z	Abundance
36.00	458.0
37.00	2633.0
38.00	2381.0
39.00	987.0
40.00	7642.0
43.90	686.0
45.00	533.0
46.20	109.0
47.00	732.0
48.00	369.0
49.00	2338.0
50.00	11888.0
51.00	3523.0
52.00	137.0
56.00	887.0
57.00	1485.0
60.00	596.0
61.10	2857.0
62.00	2934.0
63.00	2635.0
64.00	406.0
65.00	682.0
66.90	211.0
68.00	6577.0
69.00	6885.0
70.00	523.0
72.00	498.0
73.00	2922.0
74.00	11713.0
75.00	34224.0
76.10	2906.0
77.00	590.0
77.80	348.0
79.00	1940.0
80.00	755.0
80.90	2031.0
81.90	539.0
85.90	151.0
87.00	3467.0
88.00	3167.0
90.90	247.0
92.00	1826.0
93.00	2916.0
94.00	8639.0
95.00	76040.0
96.00	4871.0
97.10	104.0
103.90	458.0
105.90	242.0
116.00	281.0
117.00	436.0
118.00	331.0
118.90	483.0
127.90	241.0
129.00	110.0
129.90	323.0
130.90	167.0
134.80	138.0
140.90	636.0
141.90	110.0
142.90	716.0
145.00	239.0
148.00	176.0
150.00	116.0
154.90	220.0
172.00	1030.0
172.90	573.0
174.00	71136.0
175.00	4460.0
176.00	68144.0
177.00	4269.0

Instrument :
MSVOA_U
ClientSampleId :
BFB37