


Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU031920\
 Data File : VU037333.D
 Acq On : 19 Mar 2020 08:50
 Operator : JC/MD
 Sample : BFB06
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFB06

Integration File: rteint.p

Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR031620WMA.M
 Title : TRACE VOA SOM01.0
 Last Update : Mon Mar 23 07:30:40 2020



AutoFind: Scans 2896, 2897, 2898; Background Corrected with Scan 2883

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.9	20293	PASS
75	95	30	80	57.0	55392	PASS
95	95	100	100	100.0	97130	PASS
96	95	5	9	6.9	6656	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	120	78.9	76594	PASS
175	174	5	9	8.0	6163	PASS
176	174	95	101	98.7	75586	PASS
177	176	5	9	6.5	4933	PASS

Scan 2883 (10.610 min): VU037333.D
BFB06

m/z	Abundance
35.90	242.0
44.00	346.0

Instrument :
MSVOA_U
ClientSampleId :
BFB06

m/z	Abundance
36.00	1190.0
37.00	5513.0
38.00	4927.0
39.00	1844.0
39.90	242.0
42.90	128.0
44.00	924.0
45.00	933.0
46.00	121.0
47.10	777.0
47.90	634.0
49.00	4917.0
50.00	19944.0
51.00	6191.0
51.90	188.0
54.90	388.0
56.00	1532.0
57.00	3075.0
60.00	1111.0
61.00	5501.0
62.00	5463.0
63.00	5238.0
64.00	523.0
65.00	753.0
66.80	201.0
68.00	10557.0
69.00	11223.0
70.00	725.0
72.10	619.0
73.00	5348.0
74.00	19088.0
75.00	54544.0
76.10	4919.0
76.90	553.0
78.00	366.0
78.90	4531.0
80.00	1228.0
81.00	4779.0
81.90	1141.0
83.00	126.0
86.00	268.0
87.00	3175.0
88.00	3357.0
90.90	417.0
92.00	3211.0
93.00	4847.0
94.00	12447.0
95.10	96152.0
96.10	6501.0
96.90	186.0
103.90	741.0
105.00	142.0
105.90	583.0
112.80	143.0
115.10	183.0
115.90	453.0
116.90	892.0
117.90	564.0
118.90	732.0
128.00	424.0
129.00	211.0
130.00	353.0
130.90	159.0
134.90	213.0
136.90	271.0
140.00	137.0
141.00	1378.0
141.90	171.0
143.00	1529.0
144.90	268.0
146.00	193.0
148.00	276.0
155.00	201.0
156.90	226.0
158.90	159.0
160.80	101.0
171.20	202.0
172.00	776.0
174.10	74008.0
175.10	5985.0
176.00	72720.0
177.10	4657.0
177.90	113.0

Instrument :
MSVOA_U
ClientSampleId :
BFB06

m/z	Abundance
36.00	1236.0
37.00	5741.0
38.00	5036.0
39.00	1959.0
39.90	187.0
44.00	939.0
45.00	911.0
45.90	145.0
47.10	738.0
47.90	707.0
49.00	5047.0
50.00	20928.0
51.00	6413.0
51.90	224.0
53.10	114.0
55.00	367.0
56.00	1710.0
57.00	3131.0
58.10	101.0
60.00	1107.0
61.00	5690.0
62.00	5867.0
63.00	5252.0
64.00	599.0
65.00	793.0
65.80	118.0
68.00	10938.0
69.00	11750.0
70.00	736.0
72.10	590.0
73.00	5586.0
74.10	19832.0
75.00	57264.0
76.10	5046.0
76.90	538.0
78.00	367.0
78.90	4772.0
80.00	1297.0
81.00	5085.0
81.90	1210.0
83.00	111.0
86.00	261.0
87.00	3371.0
88.00	3437.0
90.90	393.0
92.10	3576.0
93.00	4973.0
94.00	12930.0
95.10	100152.0
96.10	6893.0
96.90	184.0
103.90	746.0
105.00	170.0
105.90	599.0
107.10	135.0
111.00	105.0
112.80	113.0
115.10	146.0
115.90	493.0
116.90	916.0
117.90	574.0
118.90	840.0
128.00	421.0
129.00	286.0
130.00	346.0
131.00	171.0
135.00	209.0
136.90	289.0
140.90	1325.0
143.00	1675.0
144.90	285.0
146.00	207.0
146.80	116.0
147.90	282.0
155.00	191.0
156.90	290.0
158.90	194.0
161.10	131.0
172.00	757.0
174.10	78968.0
175.10	6319.0
176.10	77592.0
177.10	5090.0
178.00	165.0

Instrument :
MSVOA_U
ClientSampleId :
BFB06

m/z	Abundance
36.00	1215.0
37.00	5486.0
38.00	4680.0
39.00	1897.0
39.90	120.0
44.00	932.0
45.00	871.0
45.90	126.0
47.00	741.0
47.90	667.0
49.00	4595.0
50.00	20008.0
51.00	5973.0
51.90	239.0
53.10	128.0
55.00	333.0
56.00	1764.0
57.00	2852.0
58.10	128.0
59.90	1083.0
61.00	5370.0
62.00	5652.0
63.00	4677.0
64.00	629.0
65.00	743.0
65.80	115.0
67.10	343.0
68.00	10603.0
69.00	11062.0
70.00	717.0
72.00	545.0
73.00	5264.0
74.10	18768.0
75.00	54368.0
76.10	4644.0
77.00	508.0
78.00	318.0
79.00	4578.0
80.00	1311.0
81.00	4863.0
81.90	1137.0
86.00	187.0
87.00	3332.0
88.00	3494.0
90.90	354.0
92.10	3579.0
93.00	4653.0
94.10	12207.0
95.10	95088.0
96.10	6576.0
97.00	152.0
104.00	699.0
105.00	175.0
105.90	595.0
107.10	165.0
111.00	119.0
115.90	540.0
116.90	795.0
117.90	546.0
118.90	885.0
128.00	392.0
129.00	273.0
129.90	350.0
131.00	163.0
135.00	198.0
136.90	280.0
140.90	1196.0
143.00	1675.0
144.90	264.0
146.00	225.0
146.80	116.0
147.90	239.0
155.00	173.0
157.00	313.0
158.90	179.0
161.10	180.0
172.00	664.0
174.10	76808.0
175.10	6186.0
176.10	76448.0
177.10	5054.0
178.00	187.0

Instrument :
MSVOA_U
ClientSampleId :
BFB06