

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031921\
 Data File : VU042724.D
 Acq On : 19 Mar 2021 16:37
 Operator : SY/MD
 Sample : M1744-06MS
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 ETZF6MS

Quant Time: Mar 20 05:51:55 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SOMUTR030221WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Mar 20 05:47:33 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.256	114	74303	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.423	117	72152	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.819	152	35145	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	26076	5.26	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	105.20%
7) Chloroethane-d5	1.919	69	22686	6.05	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	121.00%
11) 1,1-Dichloroethene-d2	2.575	63	53558	4.79	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	95.80%
20) 2-Butanone-d5	4.646	46	85045	52.29	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	104.58%
24) Chloroform-d	5.073	84	48417	4.94	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.80%
26) 1,2-Dichloroethane-d4	5.710	65	31289	5.19	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	103.80%
32) Benzene-d6	5.735	84	88729	5.02	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.40%
36) 1,2-Dichloropropane-d6	6.700	67	28589	5.01	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	100.20%
41) Toluene-d8	7.906	98	82531	4.78	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.60%
43) trans-1,3-Dichloroprop...	8.189	79	11166	4.19	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	83.80%
46) 2-Hexanone-d5	8.645	63	61236	47.97	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	95.94%
57) 1,1,2,2-Tetrachloroeth...	10.764	84	23700	4.66	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	93.20%
64) 1,2-Dichlorobenzene-d4	12.201	152	32549	5.11	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	102.20%
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.388	85	53863	7.79	ug/L	99
12) 1,1-Dichloroethene	2.588	96	23359	4.46	ug/L	88
14) Carbon disulfide	2.800	76	197381	11.84	ug/L	100
22) cis-1,2-Dichloroethene	4.678	96	34749	5.91	ug/L	97
33) Benzene	5.784	78	115316	5.11	ug/L	100
34) Trichloroethene	6.549	95	96345	15.39	ug/L	97
42) Toluene	7.977	91	115533	4.69	ug/L	97
47) Tetrachloroethene	8.559	164	36032	7.06	ug/L	97
51) Chlorobenzene	9.452	112	73596	4.68	ug/L	97
55) Styrene	10.121	104	779	0.05	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031921\
Data File : VU042724.D
Acq On : 19 Mar 2021 16:37
Operator : SY/MD
Sample : M1744-06MS
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
ETZF6MS

Quant Time: Mar 20 05:51:55 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SOMUTR030221WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Mar 20 05:47:33 2021
Response via : Initial Calibration

