

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031921\
 Data File : VU042727.D
 Acq On : 19 Mar 2021 17:49
 Operator : SY/MD
 Sample : M1743-03
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 ETZD4

Quant Time: Mar 20 05:53:10 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SOMUTR030221WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Mar 20 05:47:33 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.256	114	71196	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.423	117	70827	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.819	152	33279	5.00	ug/L	0.00

System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	7759	1.63	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	32.60%#
7) Chloroethane-d5	1.919	69	9199	2.56	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	51.20%#
11) 1,1-Dichloroethene-d2	2.575	63	11888	1.11	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	22.20%#
20) 2-Butanone-d5	4.652	46	68664	44.06	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	88.12%
24) Chloroform-d	5.073	84	31231	3.33	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	66.60%#
26) 1,2-Dichloroethane-d4	5.713	65	21552	3.73	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	74.60%
32) Benzene-d6	5.739	84	39059	2.25	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	45.00%#
36) 1,2-Dichloropropane-d6	6.700	67	17473	3.12	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	62.40%
41) Toluene-d8	7.906	98	27900	1.65	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	33.00%#
43) trans-1,3-Dichloroprop...	8.189	79	6627	2.53	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	50.60%#
46) 2-Hexanone-d5	8.645	63	61904	49.40	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	98.80%
57) 1,1,2,2-Tetrachloroeth...	10.764	84	23343	4.68	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	93.60%
64) 1,2-Dichlorobenzene-d4	12.201	152	21596	3.58	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	71.60%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031921\
Data File : VU042727.D
Acq On : 19 Mar 2021 17:49
Operator : SY/MD
Sample : M1743-03
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
ETZD4

Quant Time: Mar 20 05:53:10 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SOMUTR030221WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Mar 20 05:47:33 2021
Response via : Initial Calibration

