

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031921\
 Data File : VU042729.D
 Acq On : 19 Mar 2021 18:37
 Operator : SY/MD
 Sample : M1743-05
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 ETZD6

Quant Time: Mar 20 05:53:43 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SOMUTR030221WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Mar 20 05:47:33 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.260	114	70433	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.423	117	69714	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.819	152	33053	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.601	65	7453	1.59	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	31.80%#
7) Chloroethane-d5	1.919	69	8786	2.47	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	49.40%#
11) 1,1-Dichloroethene-d2	2.575	63	11708	1.10	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	22.00%#
20) 2-Butanone-d5	4.649	46	67536	43.81	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	87.62%
24) Chloroform-d	5.073	84	31657	3.41	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	68.20%#
26) 1,2-Dichloroethane-d4	5.710	65	21064	3.69	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	73.80%
32) Benzene-d6	5.735	84	39466	2.31	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	46.20%#
36) 1,2-Dichloropropane-d6	6.700	67	18109	3.28	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	65.60%
41) Toluene-d8	7.906	98	28164	1.69	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	33.80%#
43) trans-1,3-Dichloroprop...	8.185	79	6670	2.59	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	51.80%#
46) 2-Hexanone-d5	8.645	63	64080	51.95	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	103.90%
57) 1,1,2,2-Tetrachloroeth...	10.764	84	23992	4.88	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	97.60%
64) 1,2-Dichlorobenzene-d4	12.201	152	22031	3.68	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	73.60%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031921\
Data File : VU042729.D
Acq On : 19 Mar 2021 18:37
Operator : SY/MD
Sample : M1743-05
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
ETZD6

Quant Time: Mar 20 05:53:43 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SOMUTR030221WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Mar 20 05:47:33 2021
Response via : Initial Calibration

