

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031924\
 Data File : VU058153.D
 Acq On : 19 Mar 2024 13:43
 Operator : MD/SY
 Sample : VU0319WBL01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK151

Quant Time: Mar 20 05:20:27 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR022924WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Mar 20 05:19:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	155029	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	145063	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	58745	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	41265	3.699	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	74.000%
7) Chloroethane-d5	1.904	69	37785	4.121	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	82.400%
11) 1,1-Dichloroethene-d2	2.560	65	18328	3.797	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	76.000%
20) 2-Butanone-d5	4.621	46	76637	41.700	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	83.400%
24) Chloroform-d	5.052	84	84342	3.998	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	80.000%
26) 1,2-Dichloroethane-d4	5.692	65	41187	4.309	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.200%
32) Benzene-d6	5.718	84	166644	3.942	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	78.800%
36) 1,2-Dichloropropane-d6	6.682	67	53388	4.154	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	83.000%
41) Toluene-d8	7.891	98	140071	3.657	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	73.200%
43) trans-1,3-Dichloroprop...	8.177	79	14239	3.222	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	64.400%
46) 2-Hexanone-d5	8.628	63	60799	40.976	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	81.960%
56) 1,1,2,2-Tetrachloroeth...	10.750	84	32190	4.085	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	81.600%
66) 1,2-Dichlorobenzene-d4	12.187	152	42025	4.344	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	86.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031924\
Data File : VU058153.D
Acq On : 19 Mar 2024 13:43
Operator : MD/SY
Sample : VU0319WBL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK151

Quant Time: Mar 20 05:20:27 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR022924WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Mar 20 05:19:57 2024
Response via : Initial Calibration

