

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031924\
 Data File : VU058160.D
 Acq On : 19 Mar 2024 17:13
 Operator : MD/SY
 Sample : VIBLK185
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VIBLK185

Quant Time: Mar 20 05:22:07 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR022924WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Mar 20 05:19:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	129104	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	121733	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	50580	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	36083	3.884	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	77.600%
7) Chloroethane-d5	1.904	69	34982	4.582	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	91.600%
11) 1,1-Dichloroethene-d2	2.560	65	17285	4.300	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	86.000%
20) 2-Butanone-d5	4.615	46	76900	50.245	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	100.500%
24) Chloroform-d	5.052	84	79765	4.540	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.800%
26) 1,2-Dichloroethane-d4	5.692	65	38900	4.887	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.800%
32) Benzene-d6	5.718	84	156703	4.417	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.400%
36) 1,2-Dichloropropane-d6	6.682	67	50728	4.704	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	94.000%
41) Toluene-d8	7.891	98	128592	4.001	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	80.000%
43) trans-1,3-Dichloroprop...	8.174	79	13268	3.578	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	71.600%
46) 2-Hexanone-d5	8.627	63	61714	49.563	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	99.120%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	31515	4.765	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	95.400%
66) 1,2-Dichlorobenzene-d4	12.190	152	40033	4.806	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	96.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031924\
Data File : VU058160.D
Acq On : 19 Mar 2024 17:13
Operator : MD/SY
Sample : VIBLK185
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK185

Quant Time: Mar 20 05:22:07 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR022924WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Mar 20 05:19:57 2024
Response via : Initial Calibration

