

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU032019\
 Data File : VU030226.D
 Acq On : 20 Mar 2019 16:43
 Operator : JC/SP
 Sample : VU0320MBL02
 Misc : 5.00g/5mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK43

Quant Time: Mar 22 01:41:34 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM031919WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Mar 22 01:27:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	483066	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	472935	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	221640	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	163108	49.93	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	99.86%
7) Chloroethane-d5	1.68	69	136526	52.27	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	104.54%
11) 1,1-Dichloroethene-d2	2.27	63	226576	38.02	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	76.04%
21) 2-Butanone-d5	4.17	46	257358	106.20	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	106.20%
24) Chloroform-d	4.65	84	303935	50.70	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.40%
26) 1,2-Dichloroethane-d4	5.31	65	184707	54.02	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	108.04%
32) Benzene-d6	5.34	84	661593	53.09	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	106.18%
36) 1,2-Dichloropropane-d6	6.33	67	207265	53.75	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	107.50%
41) Toluene-d8	7.56	98	593697	50.66	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	101.32%
43) trans-1,3-Dichloropropene-	7.85	79	88294	47.02	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	94.04%
47) 2-Hexanone-d5	8.31	63	215774	107.90	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	107.90%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	308233	52.11	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	104.22%
64) 1,2-Dichlorobenzene-d4	11.86	152	240537	54.05	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	108.10%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU032019\
Data File : VU030226.D
Acq On : 20 Mar 2019 16:43
Operator : JC/SP
Sample : VU0320MBL02
Misc : 5.00µ/5mL/100µL/5.0mL/MSVOA_U/MEOH
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK43

Quant Time: Mar 22 01:41:34 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM031919WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Mar 22 01:27:23 2019
Response via : Initial Calibration

