

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU032020\
 Data File : VU037357.D
 Acq On : 20 Mar 2020 20:36
 Operator : JC/MD
 Sample : L1983-06
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK01

Quant Time: Mar 21 01:00:31 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR032020WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Mar 21 00:40:26 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.28	114	145032	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.44	117	141089	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.83	152	62744	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	45767	4.77	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	95.40%
7) Chloroethane-d5	1.93	69	37894	4.80	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	96.00%
11) 1,1-Dichloroethene-d2	2.59	63	62124	3.30	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	66.00%
20) 2-Butanone-d5	4.68	46	114810	50.30	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	100.60%
24) Chloroform-d	5.10	84	93607	4.78	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.60%
26) 1,2-Dichloroethane-d4	5.74	65	54836	4.82	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.40%
32) Benzene-d6	5.76	84	170128	4.97	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.40%
36) 1,2-Dichloropropane-d6	6.72	67	54102	5.01	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	100.20%
41) Toluene-d8	7.93	98	155307	4.57	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.40%
43) trans-1,3-Dichloropropene-	8.20	79	23695	4.81	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	96.20%
46) 2-Hexanone-d5	8.66	63	93094	50.58	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	101.16%
57) 1,1,2,2-Tetrachloroethane-	10.77	84	43440	4.81	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	96.20%
64) 1,2-Dichlorobenzene-d4	12.21	152	57698	5.01	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	100.20%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU032020\
Data File : VU037357.D
Acq On : 20 Mar 2020 20:36
Operator : JC/MD
Sample : L1983-06
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK01

Quant Time: Mar 21 01:00:31 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR032020WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Mar 21 00:40:26 2020
Response via : Initial Calibration

