

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU032024\
 Data File : VU058176.D
 Acq On : 20 Mar 2024 16:34
 Operator : MD/SY
 Sample : P1779-08DL 100X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AK2DL

Quant Time: Mar 21 01:50:05 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR022924WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Mar 21 01:49:11 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	108544	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	104571	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	46622	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	30360	3.887	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	77.800%	
7) Chloroethane-d5	1.911	69	29816	4.645	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	93.000%	
11) 1,1-Dichloroethene-d2	2.563	65	13693	4.052	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	81.000%	
20) 2-Butanone-d5	4.628	46	79121	61.489	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	122.980%	
24) Chloroform-d	5.052	84	69751	4.722	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	94.400%	
26) 1,2-Dichloroethane-d4	5.695	65	35334	5.280	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	105.600%	
32) Benzene-d6	5.721	84	132406	4.345	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	86.800%	
36) 1,2-Dichloropropane-d6	6.682	67	43393	4.684	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	93.600%	
41) Toluene-d8	7.891	98	110800	4.013	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	80.200%	
43) trans-1,3-Dichloroprop...	8.174	79	12509	3.927	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	78.600%	
46) 2-Hexanone-d5	8.631	63	61747	57.729	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	115.460%	
56) 1,1,2,2-Tetrachloroeth...	10.750	84	30529	5.374	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	107.400%	
66) 1,2-Dichlorobenzene-d4	12.190	152	37277	4.855	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	97.000%	
Target Compounds						
12) 1,1-Dichloroethene	2.576	96	1258	0.192	ug/L	1
16) Methylene chloride	3.039	84	3166	0.398	ug/L	91
22) cis-1,2-Dichloroethene	4.666	96	5493	0.700	ug/L	94
34) Trichloroethene	6.538	95	58112	6.883	ug/L	96
47) Tetrachloroethene	8.550	164	3636	0.621	ug/L	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU032024\
Data File : VU058176.D
Acq On : 20 Mar 2024 16:34
Operator : MD/SY
Sample : P1779-08DL 100X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AK2DL

Quant Time: Mar 21 01:50:05 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR022924WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Mar 21 01:49:11 2024
Response via : Initial Calibration

