

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU032123\
 Data File : VU053315.D
 Acq On : 21 Mar 2023 11:39
 Operator : JC/MD
 Sample : VU0321WBL01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK055


Quant Time: Mar 22 03:09:19 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM031523WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Mar 22 03:08:43 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.243	114	117052	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.414	117	106931	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.806	152	46384	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	47199	43.708	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	87.420%
7) Chloroethane-d5	1.912	69	44366	50.792	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	101.580%
11) 1,1-Dichloroethene-d2	2.565	63	61086	32.654	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	65.300%
21) 2-Butanone-d5	4.623	46	53663	107.144	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	107.140%
24) Chloroform-d	5.057	84	79119	44.484	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.960%
26) 1,2-Dichloroethane-d4	5.697	65	54507	49.346	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.700%
32) Benzene-d6	5.723	84	165845	47.817	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.640%
36) 1,2-Dichloropropane-d6	6.684	67	55055	49.183	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	98.360%
41) Toluene-d8	7.893	98	151793	47.007	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.020%
43) trans-1,3-Dichloroprop...	8.176	79	25551	46.264	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	92.520%
47) 2-Hexanone-d5	8.626	63	36216	103.485	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	103.490%
56) 1,1,2,2-Tetrachloroeth...	10.751	84	73344	50.579	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	101.160%
66) 1,2-Dichlorobenzene-d4	12.189	152	46611	51.898	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	103.800%
Target Compounds						
					Qvalue	
10) 1,1,2-Trichloro-1,2,2-...	2.562	101	583	0.627	ug/L #	22
33) Benzene	5.771	78	4590	1.238	ug/L	100
35) Methylcyclohexane	6.684	83	11791	6.959	ug/L #	21
37) 1,2-Dichloropropane	6.684	63	6089	5.887	ug/L #	90
44) trans-1,3-Dichloropropene	8.176	75	1336	0.860	ug/L	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Mar 22 03:09:19 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM031523WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Mar 22 03:08:43 2023
Response via : Initial Calibration









#10
 1,1,2-Trichloro-1,2,2-trifluoroethane
 Concen: 0.627 ug/L
 RT: 2.562 min Scan# 386
 Delta R.T. -0.019 min
 Lab File: VU053315.D
 Acq: 21 Mar 2023 11:39




Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK055


Tgt Ion:101 Resp: 583
 Ion Ratio Lower Upper
 101 100
 85 0.0 37.0 55.6#
 151 0.0 52.9 79.3#



#33
 Benzene
 Concen: 1.238 ug/L
 RT: 5.771 min Scan# 1378
 Delta R.T. 0.000 min
 Lab File: VU053315.D
 Acq: 21 Mar 2023 11:39

Tgt Ion: 78 Resp: 4590





#35

Methylcyclohexane

Concen: 6.959 ug/L

RT: 6.684 min Scan# 1662

Delta R.T. -0.074 min

Lab File: VU053315.D

Acq: 21 Mar 2023 11:39

Instrument :

MSVOA_U

ClientSampleId :

VBLK055





Tgt Ion: 83 Resp: 11791

Ion Ratio Lower Upper

83 100

55 0.0 58.2 87.4#

98 0.6 35.4 53.2#



#37

1,2-Dichloropropane

Concen: 5.887 ug/L

RT: 6.684 min Scan# 1662

Delta R.T. -0.103 min

Lab File: VU053315.D




Acq: 21 Mar 2023 11:39


Tgt Ion: 63 Resp: 6089

Ion Ratio Lower Upper

63 100

112 0.0 2.6 3.8#





#44
 trans-1,3-Dichloropropene
 Concen: 0.860 ug/L
 RT: 8.176 min Scan# 21
 Delta R.T. -0.029 min
 Lab File: VU053315.D
 Acq: 21 Mar 2023 11:39

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK055

Tgt Ion: 75 Resp: 1336
 Ion Ratio Lower Upper
 75 100
 77 37.1 22.0 40.8

