

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU032219\
 Data File : VU030304.D
 Acq On : 22 Mar 2019 13:36
 Operator : JC/SP
 Sample : K2048-07 50X
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BF199

Quant Time: Mar 23 04:13:38 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM031919WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Mar 23 02:31:55 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	519575	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	502734	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	226695	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	151839	43.22	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	86.44%
7) Chloroethane-d5	1.68	69	130110	46.31	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	92.62%
11) 1,1-Dichloroethene-d2	2.27	63	212797	33.20	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	66.40%
21) 2-Butanone-d5	4.17	46	262576	100.74	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	100.74%
24) Chloroform-d	4.65	84	297072	46.07	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.14%
26) 1,2-Dichloroethane-d4	5.31	65	175382	47.69	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.38%
32) Benzene-d6	5.34	84	616528	46.54	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.08%
36) 1,2-Dichloropropane-d6	6.33	67	203254	49.58	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	99.16%
41) Toluene-d8	7.56	98	556357	44.66	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	89.32%
43) trans-1,3-Dichloropropene-	7.85	79	87727	43.95	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	87.90%
47) 2-Hexanone-d5	8.31	63	212411	99.92	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	99.92%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	304653	48.45	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	96.90%
64) 1,2-Dichlorobenzene-d4	11.86	152	226568	49.78	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.56%

Target Compounds					Ovalue
20) cis-1,2-Dichloroethene	4.23	96	10821	2.704 ug/L	88
34) Trichloroethene	6.18	95	404363	105.771 ug/L	97
46) Tetrachloroethene	8.23	164	3427	1.061 ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU032219\
Data File : VU030304.D
Acq On : 22 Mar 2019 13:36
Operator : JC/SP
Sample : K2048-07 50X
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BF199

Quant Time: Mar 23 04:13:38 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM031919WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Mar 23 02:31:55 2019
Response via : Initial Calibration

