

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU032219\
 Data File : VU030331.D
 Acq On : 23 Mar 2019 00:04
 Operator : JC/SP
 Sample : K2062-05
 Misc : 5.93g/5mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BF0H2

Quant Time: Mar 23 05:53:56 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM031919WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Mar 23 04:35:39 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	715672	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	718863	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.49	152	372388	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	115133	23.79	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	47.58%#
7) Chloroethane-d5	1.64	69	57108	14.76	ug/L	-0.04
Spiked Amount	50.000	Range	70 - 130	Recovery	=	29.52%#
11) 1,1-Dichloroethene-d2	2.22	63	166688	18.88	ug/L	-0.05
Spiked Amount	50.000	Range	60 - 125	Recovery	=	37.76%#
21) 2-Butanone-d5	4.20	46	251345	70.01	ug/L	0.02
Spiked Amount	100.000	Range	40 - 130	Recovery	=	70.01%
24) Chloroform-d	4.64	84	259962	29.27	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	58.54%#
26) 1,2-Dichloroethane-d4	5.30	65	157202	31.03	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	62.06%#
32) Benzene-d6	5.33	84	526605	27.80	ug/L	-0.01
Spiked Amount	50.000	Range	70 - 125	Recovery	=	55.60%#
36) 1,2-Dichloropropane-d6	6.32	67	175992	30.03	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	60.06%#
41) Toluene-d8	7.56	98	460859	25.87	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	51.74%#
43) trans-1,3-Dichloropropene-	7.85	79	76813	26.91	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	53.82%#
47) 2-Hexanone-d5	8.32	63	210531	69.26	ug/L	0.02
Spiked Amount	100.000	Range	45 - 130	Recovery	=	69.26%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	312983	34.81	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	69.62%
64) 1,2-Dichlorobenzene-d4	11.86	152	225829	30.21	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	60.42%#

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU032219\
Data File : VU030331.D
Acq On : 23 Mar 2019 00:04
Operator : JC/SP
Sample : K2062-05
Misc : 5.93q/5mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BF0H2

Quant Time: Mar 23 05:53:56 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM031919WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Mar 23 04:35:39 2019
Response via : Initial Calibration

