

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU032219\
 Data File : VU030332.D
 Acq On : 23 Mar 2019 00:27
 Operator : JC/SP
 Sample : K2062-11
 Misc : 6.16g/5mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BF1H7

Quant Time: Mar 23 05:55:39 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM031919WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Mar 23 04:35:39 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	539318	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	550388	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	279424	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	118516	32.50	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	65.00%
7) Chloroethane-d5	1.64	69	57508	19.72	ug/L	-0.04
Spiked Amount	50.000	Range	70 - 130	Recovery	=	39.44%#
11) 1,1-Dichloroethene-d2	2.22	63	169476	25.47	ug/L	-0.05
Spiked Amount	50.000	Range	60 - 125	Recovery	=	50.94%#
21) 2-Butanone-d5	4.19	46	256073	94.65	ug/L	0.02
Spiked Amount	100.000	Range	40 - 130	Recovery	=	94.65%
24) Chloroform-d	4.64	84	261227	39.03	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	78.06%
26) 1,2-Dichloroethane-d4	5.30	65	159121	41.68	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	83.36%
32) Benzene-d6	5.33	84	537507	37.06	ug/L	-0.01
Spiked Amount	50.000	Range	70 - 125	Recovery	=	74.12%
36) 1,2-Dichloropropane-d6	6.32	67	178867	39.86	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	79.72%
41) Toluene-d8	7.56	98	470518	34.50	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	69.00%#
43) trans-1,3-Dichloropropene-	7.85	79	80097	36.65	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	73.30%
47) 2-Hexanone-d5	8.32	63	216676	93.10	ug/L	0.02
Spiked Amount	100.000	Range	45 - 130	Recovery	=	93.10%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	322774	46.89	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	93.78%
64) 1,2-Dichlorobenzene-d4	11.86	152	230677	41.12	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	82.24%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU032219\
Data File : VU030332.D
Acq On : 23 Mar 2019 00:27
Operator : JC/SP
Sample : K2062-11
Misc : 6.16g/5mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BF1H7

Quant Time: Mar 23 05:55:39 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM031919WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Mar 23 04:35:39 2019
Response via : Initial Calibration

