

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU032221\
 Data File : VU042754.D
 Acq On : 22 Mar 2021 11:51
 Operator : SY/MD
 Sample : M1743-02
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 ETZD3

Quant Time: Mar 23 07:17:15 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SOMUTR030221WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Mar 23 07:16:03 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.260	114	72300	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.427	117	73432	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.819	152	34403	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	25337	5.26	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	105.20%
7) Chloroethane-d5	1.919	69	20477	5.61	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	112.20%
11) 1,1-Dichloroethene-d2	2.575	63	39787	3.66	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	73.20%
20) 2-Butanone-d5	4.649	46	87376	55.21	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	110.42%
24) Chloroform-d	5.073	84	48825	5.12	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.40%
26) 1,2-Dichloroethane-d4	5.713	65	31278	5.33	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	106.60%
32) Benzene-d6	5.736	84	89241	4.96	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.20%
36) 1,2-Dichloropropane-d6	6.700	67	29091	5.01	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	100.20%
41) Toluene-d8	7.906	98	81859	4.66	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.20%
43) trans-1,3-Dichloroprop...	8.189	79	11922	4.39	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	87.80%
46) 2-Hexanone-d5	8.645	63	61964	47.69	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	95.38%
57) 1,1,2,2-Tetrachloroeth...	10.764	84	24984	4.83	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	96.60%
64) 1,2-Dichlorobenzene-d4	12.201	152	32282	5.18	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	103.60%
Target Compounds						Qvalue
25) Chloroform	5.102	83	2485	0.23	ug/L	95
47) Tetrachloroethene	8.562	164	11594	2.23	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU032221\
Data File : VU042754.D
Acq On : 22 Mar 2021 11:51
Operator : SY/MD
Sample : M1743-02
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
ETZD3

Quant Time: Mar 23 07:17:15 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SOMUTR030221WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Mar 23 07:16:03 2021
Response via : Initial Calibration

