

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_U\Data\VU032221\  
 Data File : VU042755.D  
 Acq On : 22 Mar 2021 12:15  
 Operator : SY/MD  
 Sample : M1743-03  
 Misc : 25.0mL/MSVOA\_U/WATER  
 ALS Vial : 7 Sample Multiplier: 1

Instrument :  
 MSVOA\_U  
 ClientSampleId :  
 ETZD4

Quant Time: Mar 23 07:17:36 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_U\Method\SOMUTR030221WMA.M  
 Quant Title : TRACE VOA SOM01.0  
 QLast Update : Tue Mar 23 07:16:03 2021  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon  | Response | Conc     | Units | Dev(Min)  |
|-------------------------------|--------|-------|----------|----------|-------|-----------|
| -----                         |        |       |          |          |       |           |
| Internal Standards            |        |       |          |          |       |           |
| 1) 1,4-Difluorobenzene        | 6.260  | 114   | 76753    | 5.00     | ug/L  | 0.00      |
| 28) Chlorobenzene-d5          | 9.427  | 117   | 76966    | 5.00     | ug/L  | 0.00      |
| 60) 1,4-Dichlorobenzene-d4    | 11.819 | 152   | 35939    | 5.00     | ug/L  | 0.00      |
| System Monitoring Compounds   |        |       |          |          |       |           |
| 4) Vinyl Chloride-d3          | 1.604  | 65    | 24444    | 4.78     | ug/L  | 0.00      |
| Spiked Amount                 | 5.000  | Range | 40 - 130 | Recovery | =     | 95.60%    |
| 7) Chloroethane-d5            | 1.919  | 69    | 19589    | 5.06     | ug/L  | 0.00      |
| Spiked Amount                 | 5.000  | Range | 65 - 130 | Recovery | =     | 101.20%   |
| 11) 1,1-Dichloroethene-d2     | 2.575  | 63    | 41737    | 3.61     | ug/L  | 0.00      |
| Spiked Amount                 | 5.000  | Range | 60 - 125 | Recovery | =     | 72.20%    |
| 20) 2-Butanone-d5             | 4.652  | 46    | 65652    | 39.08    | ug/L  | 0.00      |
| Spiked Amount                 | 50.000 | Range | 40 - 130 | Recovery | =     | 78.16%    |
| 24) Chloroform-d              | 5.073  | 84    | 49913    | 4.93     | ug/L  | 0.00      |
| Spiked Amount                 | 5.000  | Range | 70 - 125 | Recovery | =     | 98.60%    |
| 26) 1,2-Dichloroethane-d4     | 5.710  | 65    | 30868    | 4.96     | ug/L  | 0.00      |
| Spiked Amount                 | 5.000  | Range | 70 - 130 | Recovery | =     | 99.20%    |
| 32) Benzene-d6                | 5.736  | 84    | 89911    | 4.76     | ug/L  | 0.00      |
| Spiked Amount                 | 5.000  | Range | 70 - 125 | Recovery | =     | 95.20%    |
| 36) 1,2-Dichloropropane-d6    | 6.700  | 67    | 29744    | 4.89     | ug/L  | 0.00      |
| Spiked Amount                 | 5.000  | Range | 60 - 140 | Recovery | =     | 97.80%    |
| 41) Toluene-d8                | 7.906  | 98    | 83180    | 4.51     | ug/L  | 0.00      |
| Spiked Amount                 | 5.000  | Range | 70 - 130 | Recovery | =     | 90.20%    |
| 43) trans-1,3-Dichloroprop... | 8.189  | 79    | 12261    | 4.31     | ug/L  | 0.00      |
| Spiked Amount                 | 5.000  | Range | 55 - 130 | Recovery | =     | 86.20%    |
| 46) 2-Hexanone-d5             | 8.645  | 63    | 60911    | 44.73    | ug/L  | 0.00      |
| Spiked Amount                 | 50.000 | Range | 45 - 130 | Recovery | =     | 89.46%    |
| 57) 1,1,2,2-Tetrachloroeth... | 10.764 | 84    | 25251    | 4.66     | ug/L  | 0.00      |
| Spiked Amount                 | 5.000  | Range | 65 - 120 | Recovery | =     | 93.20%    |
| 64) 1,2-Dichlorobenzene-d4    | 12.201 | 152   | 32495    | 4.99     | ug/L  | 0.00      |
| Spiked Amount                 | 5.000  | Range | 80 - 120 | Recovery | =     | 99.80%    |
| Target Compounds              |        |       |          |          |       |           |
| 42) Toluene                   | 7.977  | 91    | 936      | 0.04     | ug/L  | Qvalue 95 |
| -----                         |        |       |          |          |       |           |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_U\Data\VU032221\  
Data File : VU042755.D  
Acq On : 22 Mar 2021 12:15  
Operator : SY/MD  
Sample : M1743-03  
Misc : 25.0mL/MSVOA\_U/WATER  
ALS Vial : 7 Sample Multiplier: 1

Instrument :  
MSVOA\_U  
ClientSampleId :  
ETZD4

Quant Time: Mar 23 07:17:36 2021  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_U\Method\SOMUTR030221WMA.M  
Quant Title : TRACE VOA SOM01.0  
QLast Update : Tue Mar 23 07:16:03 2021  
Response via : Initial Calibration

