

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU032223\
 Data File : VU053363.D
 Acq On : 22 Mar 2023 19:31
 Operator : JC/MD
 Sample : 01961-05
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Quant Time: Mar 23 03:56:59 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM031523WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Mar 23 02:41:20 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.243	114	104405	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.414	117	96880	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.809	152	45542	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	39590	41.103	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	82.200%	
7) Chloroethane-d5	1.912	69	40299	51.725	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	103.440%	
11) 1,1-Dichloroethene-d2	2.565	63	52739	31.608	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	63.220%	
21) 2-Butanone-d5	4.620	46	54667	122.370	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	122.370%	
24) Chloroform-d	5.057	84	69818	44.009	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	88.020%	
26) 1,2-Dichloroethane-d4	5.697	65	49714	50.459	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	100.920%	
32) Benzene-d6	5.722	84	148998	47.416	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	94.840%	
36) 1,2-Dichloropropane-d6	6.684	67	50125	49.424	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	98.840%	
41) Toluene-d8	7.893	98	133125	45.503	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	91.000%	
43) trans-1,3-Dichloroprop...	8.176	79	22590	45.146	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	90.300%	
47) 2-Hexanone-d5	8.629	63	35017	110.440	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	110.440%	
56) 1,1,2,2-Tetrachloroeth...	10.751	84	66846	50.880	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	101.760%	
66) 1,2-Dichlorobenzene-d4	12.188	152	44029	49.930	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	99.860%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU032223\
Data File : VU053363.D
Acq On : 22 Mar 2023 19:31
Operator : JC/MD
Sample : 01961-05
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Quant Time: Mar 23 03:56:59 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM031523WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Mar 23 02:41:20 2023
Response via : Initial Calibration

