

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU032223\
 Data File : VU053345.D
 Acq On : 22 Mar 2023 12:27
 Operator : JC/MD
 Sample : 01961-04DL 10X
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 EYFG8DL

Quant Time: Mar 23 02:43:04 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM031523WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Mar 23 02:41:20 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	6.244	114	112526	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.411	117	105365	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.806	152	48692	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	41153	39.642	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	79.280%		
7) Chloroethane-d5	1.912	69	40155	47.820	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	95.640%		
11) 1,1-Dichloroethene-d2	2.565	63	53570	29.789	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	59.580%#		
21) 2-Butanone-d5	4.623	46	51483	106.926	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	106.930%		
24) Chloroform-d	5.057	84	72118	42.178	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	84.360%		
26) 1,2-Dichloroethane-d4	5.697	65	50038	47.122	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	94.240%		
32) Benzene-d6	5.723	84	151441	44.313	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	88.620%		
36) 1,2-Dichloropropane-d6	6.684	67	50523	45.805	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	91.620%		
41) Toluene-d8	7.893	98	138175	43.426	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	86.860%		
43) trans-1,3-Dichloroprop...	8.173	79	22996	42.256	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	84.520%		
47) 2-Hexanone-d5	8.629	63	32784	95.071	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	95.070%		
56) 1,1,2,2-Tetrachloroeth...	10.751	84	67517	47.253	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	94.500%		
66) 1,2-Dichlorobenzene-d4	12.189	152	43934	46.599	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	93.200%		
Target Compounds						Qvalue
20) cis-1,2-Dichloroethene	4.662	96	57059	57.938	ug/L	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU032223\
Data File : VU053345.D
Acq On : 22 Mar 2023 12:27
Operator : JC/MD
Sample : 01961-04DL 10X
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
EYFG8DL


Quant Time: Mar 23 02:43:04 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM031523WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Mar 23 02:41:20 2023
Response via : Initial Calibration





#20
 cis-1,2-Dichloroethene
 Concen: 57.938 ug/L
 RT: 4.662 min Scan# 10
 Delta R.T. 0.000 min
 Lab File: VU053345.D
 Acq: 22 Mar 2023 12:27

Instrument :
 MSVOA_U
 ClientSampleId :
 EYFG8DL



Tgt Ion:	96	Resp:	57059
Ion	Ratio	Lower	Upper
96	100		
61	130.4	92.5	171.7
68	0.0	0.0	0.0

