

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_U\Data\VU032223\  
 Data File : VU053361.D  
 Acq On : 22 Mar 2023 18:44  
 Operator : JC/MD  
 Sample : 01995-05  
 Misc : 5.0mL/MSVOA\_U/WATER  
 ALS Vial : 24 Sample Multiplier: 1

Instrument :  
 MSVOA\_U  
 ClientSampleId :  
 VHBLK001

Quant Time: Mar 23 03:56:34 2023  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_U\Method\SFAMULM031523WMA.M  
 Quant Title : VOC Analysis  
 QLast Update : Thu Mar 23 02:41:20 2023  
 Response via : Initial Calibration

| Compound                      | R.T.    | QIon           | Response | Conc    | Units    | Dev(Min) |
|-------------------------------|---------|----------------|----------|---------|----------|----------|
| -----                         |         |                |          |         |          |          |
| Internal Standards            |         |                |          |         |          |          |
| 1) 1,4-Difluorobenzene        | 6.243   | 114            | 102084   | 50.000  | ug/L     | 0.00     |
| 28) Chlorobenzene-d5          | 9.414   | 117            | 97872    | 50.000  | ug/L     | 0.00     |
| 58) 1,4-Dichlorobenzene-d4    | 11.809  | 152            | 45558    | 50.000  | ug/L     | 0.00     |
| System Monitoring Compounds   |         |                |          |         |          |          |
| 4) Vinyl Chloride-d3          | 1.597   | 65             | 39782    | 42.241  | ug/L     | 0.00     |
| Spiked Amount                 | 50.000  | Range 60 - 135 | Recovery | =       | 84.480%  |          |
| 7) Chloroethane-d5            | 1.912   | 69             | 39041    | 51.249  | ug/L     | 0.00     |
| Spiked Amount                 | 50.000  | Range 70 - 130 | Recovery | =       | 102.500% |          |
| 11) 1,1-Dichloroethene-d2     | 2.565   | 63             | 52072    | 31.917  | ug/L     | 0.00     |
| Spiked Amount                 | 50.000  | Range 60 - 125 | Recovery | =       | 63.840%  |          |
| 21) 2-Butanone-d5             | 4.620   | 46             | 53020    | 121.382 | ug/L     | 0.00     |
| Spiked Amount                 | 100.000 | Range 40 - 130 | Recovery | =       | 121.380% |          |
| 24) Chloroform-d              | 5.057   | 84             | 69979    | 45.114  | ug/L     | 0.00     |
| Spiked Amount                 | 50.000  | Range 70 - 125 | Recovery | =       | 90.220%  |          |
| 26) 1,2-Dichloroethane-d4     | 5.697   | 65             | 48004    | 49.831  | ug/L     | 0.00     |
| Spiked Amount                 | 50.000  | Range 70 - 125 | Recovery | =       | 99.660%  |          |
| 32) Benzene-d6                | 5.722   | 84             | 148153   | 46.670  | ug/L     | 0.00     |
| Spiked Amount                 | 50.000  | Range 70 - 125 | Recovery | =       | 93.340%  |          |
| 36) 1,2-Dichloropropane-d6    | 6.684   | 67             | 50081    | 48.881  | ug/L     | 0.00     |
| Spiked Amount                 | 50.000  | Range 70 - 120 | Recovery | =       | 97.760%  |          |
| 41) Toluene-d8                | 7.893   | 98             | 133207   | 45.070  | ug/L     | 0.00     |
| Spiked Amount                 | 50.000  | Range 80 - 120 | Recovery | =       | 90.140%  |          |
| 43) trans-1,3-Dichloroprop... | 8.176   | 79             | 22516    | 44.542  | ug/L     | 0.00     |
| Spiked Amount                 | 50.000  | Range 60 - 125 | Recovery | =       | 89.080%  |          |
| 47) 2-Hexanone-d5             | 8.629   | 63             | 34967    | 109.164 | ug/L     | 0.00     |
| Spiked Amount                 | 100.000 | Range 45 - 130 | Recovery | =       | 109.160% |          |
| 56) 1,1,2,2-Tetrachloroeth... | 10.751  | 84             | 68877    | 51.895  | ug/L     | 0.00     |
| Spiked Amount                 | 50.000  | Range 65 - 120 | Recovery | =       | 103.780% |          |
| 66) 1,2-Dichlorobenzene-d4    | 12.188  | 152            | 43053    | 48.806  | ug/L     | 0.00     |
| Spiked Amount                 | 50.000  | Range 80 - 120 | Recovery | =       | 97.620%  |          |

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_U\Data\VU032223\  
Data File : VU053361.D  
Acq On : 22 Mar 2023 18:44  
Operator : JC/MD  
Sample : 01995-05  
Misc : 5.0mL/MSVOA\_U/WATER  
ALS Vial : 24 Sample Multiplier: 1

Instrument :  
MSVOA\_U  
ClientSampleId :  
VHBLK001

Quant Time: Mar 23 03:56:34 2023  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_U\Method\SFAMULM031523WMA.M  
Quant Title : VOC Analysis  
QLast Update : Thu Mar 23 02:41:20 2023  
Response via : Initial Calibration

