

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU032224\
 Data File : VU058190.D
 Acq On : 22 Mar 2024 13:41
 Operator : MD/SY
 Sample : P1779-07
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AK1

Quant Time: Mar 22 23:44:17 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR032124WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Mar 22 23:43:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	94140	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	91781	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	39355	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	32973	4.145	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	82.800%	
7) Chloroethane-d5	1.911	69	29420	4.531	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	90.600%	
11) 1,1-Dichloroethene-d2	2.563	65	14334	4.340	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	86.800%	
20) 2-Butanone-d5	4.634	46	68784	48.668	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	97.340%	
24) Chloroform-d	5.055	84	63320	4.452	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	89.000%	
26) 1,2-Dichloroethane-d4	5.695	65	30700	4.617	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	92.400%	
32) Benzene-d6	5.721	84	128644	4.315	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	86.400%	
36) 1,2-Dichloropropane-d6	6.682	67	39919	4.544	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	90.800%	
41) Toluene-d8	7.891	98	108546	4.110	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	82.200%	
43) trans-1,3-Dichloroprop...	8.174	79	12220	4.341	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	86.800%	
46) 2-Hexanone-d5	8.631	63	48039	42.137	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	84.280%	
56) 1,1,2,2-Tetrachloroeth...	10.750	84	24326	4.250	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	85.000%	
66) 1,2-Dichlorobenzene-d4	12.187	152	33895	4.625	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	92.400%	
Target Compounds					Qvalue	
13) Acetone	2.660	43	24345	26.615	ug/L	97
21) 2-Butanone	4.721	43	26909	19.364	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU032224\
Data File : VU058190.D
Acq On : 22 Mar 2024 13:41
Operator : MD/SY
Sample : P1779-07
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AK1

Quant Time: Mar 22 23:44:17 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR032124WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Mar 22 23:43:24 2024
Response via : Initial Calibration

