

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU032423\
 Data File : VU053396.D
 Acq On : 24 Mar 2023 17:36
 Operator : JC/MD
 Sample : 02041-03
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 CB958

Quant Time: Mar 25 00:40:46 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR032423WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Mar 25 00:39:08 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.243	114	144595	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.414	117	133872	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.809	152	63582	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	46512	4.676	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	93.600%
7) Chloroethane-d5	1.912	69	42859	4.747	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	95.000%
11) 1,1-Dichloroethene-d2	2.565	65	19088	4.505	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	90.000%
20) 2-Butanone-d5	4.632	46	113619	49.860	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	99.720%
24) Chloroform-d	5.057	84	87306	4.734	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.600%
26) 1,2-Dichloroethane-d4	5.697	65	45910	4.858	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.200%
32) Benzene-d6	5.722	84	186109	4.905	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.200%
36) 1,2-Dichloropropane-d6	6.687	67	59223	5.123	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	102.400%
41) Toluene-d8	7.893	98	169308	4.730	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.600%
43) trans-1,3-Dichloroprop...	8.176	79	20163	4.545	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	90.800%
46) 2-Hexanone-d5	8.629	63	88465	49.094	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	98.180%
56) 1,1,2,2-Tetrachloroeth...	10.751	84	45287	4.706	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	94.200%
66) 1,2-Dichlorobenzene-d4	12.188	152	58986	5.157	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	103.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU032423\
Data File : VU053396.D
Acq On : 24 Mar 2023 17:36
Operator : JC/MD
Sample : 02041-03
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
CB958

Quant Time: Mar 25 00:40:46 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR032423WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Mar 25 00:39:08 2023
Response via : Initial Calibration

