

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU032423\
 Data File : VU053397.D
 Acq On : 24 Mar 2023 18:00
 Operator : JC/MD
 Sample : 02041-02
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 CB956

Quant Time: Mar 25 00:40:59 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR032423WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Mar 25 00:39:08 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.243	114	138012	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.414	117	129065	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.809	152	62578	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	49088	5.170	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	= 103.400%	
7) Chloroethane-d5	1.912	69	44993	5.221	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	= 104.400%	
11) 1,1-Dichloroethene-d2	2.565	65	21548	5.328	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	= 106.600%	
20) 2-Butanone-d5	4.639	46	127274	58.516	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	= 117.040%	
24) Chloroform-d	5.057	84	93111	5.290	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	= 105.800%	
26) 1,2-Dichloroethane-d4	5.697	65	47231	5.236	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	= 104.800%	
32) Benzene-d6	5.722	84	195366	5.341	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	= 106.800%	
36) 1,2-Dichloropropane-d6	6.684	67	60317	5.412	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	= 108.200%	
41) Toluene-d8	7.893	98	178623	5.176	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	= 103.600%	
43) trans-1,3-Dichloroprop...	8.176	79	22088	5.164	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	= 103.200%	
46) 2-Hexanone-d5	8.629	63	95563	55.008	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	= 110.020%	
56) 1,1,2,2-Tetrachloroeth...	10.751	84	49274	5.311	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	= 106.200%	
66) 1,2-Dichlorobenzene-d4	12.192	152	62797	5.578	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	= 111.600%	
Target Compounds						
17) Methyl tert-butyl Ether	3.359	73	4745	0.250	ug/L	# 91
51) Chlorobenzene	9.446	112	6593	0.277	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU032423\
Data File : VU053397.D
Acq On : 24 Mar 2023 18:00
Operator : JC/MD
Sample : 02041-02
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
CB956

Quant Time: Mar 25 00:40:59 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR032423WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Mar 25 00:39:08 2023
Response via : Initial Calibration

