

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU032423\
 Data File : VU053399.D
 Acq On : 24 Mar 2023 18:48
 Operator : JC/MD
 Sample : 02041-04
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Quant Time: Mar 25 00:41:36 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR032423WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Mar 25 00:39:08 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.247	114	146083	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.414	117	135173	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.806	152	64590	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	51067	5.082	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	= 101.600%		
7) Chloroethane-d5	1.912	69	47480	5.206	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	= 104.200%		
11) 1,1-Dichloroethene-d2	2.568	65	21477	5.017	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	= 100.400%		
20) 2-Butanone-d5	4.633	46	122505	53.211	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	= 106.420%		
24) Chloroform-d	5.057	84	96743	5.193	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 103.800%		
26) 1,2-Dichloroethane-d4	5.697	65	49074	5.140	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 102.800%		
32) Benzene-d6	5.723	84	199707	5.213	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 104.200%		
36) 1,2-Dichloropropane-d6	6.684	67	63265	5.420	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	= 108.400%		
41) Toluene-d8	7.893	98	188580	5.218	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 104.400%		
43) trans-1,3-Dichloroprop...	8.176	79	22497	5.022	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	= 100.400%		
46) 2-Hexanone-d5	8.629	63	98205	53.975	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	= 107.940%		
56) 1,1,2,2-Tetrachloroeth...	10.751	84	49229	5.067	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	= 101.400%		
66) 1,2-Dichlorobenzene-d4	12.189	152	61911	5.328	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	= 106.600%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU032423\
Data File : VU053399.D
Acq On : 24 Mar 2023 18:48
Operator : JC/MD
Sample : 02041-04
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Quant Time: Mar 25 00:41:36 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR032423WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Mar 25 00:39:08 2023
Response via : Initial Calibration

