

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU032423\
 Data File : VU053405.D
 Acq On : 24 Mar 2023 21:12
 Operator : JC/MD
 Sample : 02053-06
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 EYF72

Quant Time: Mar 25 00:42:57 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR032423WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Mar 25 00:39:08 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.243	114	142902	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.414	117	133640	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.806	152	63487	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	46646	4.745	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	94.800%
7) Chloroethane-d5	1.912	69	45617	5.113	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	102.200%
11) 1,1-Dichloroethene-d2	2.565	65	21013	5.018	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	100.400%
20) 2-Butanone-d5	4.633	46	120144	53.348	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	106.700%
24) Chloroform-d	5.057	84	92479	5.074	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.400%
26) 1,2-Dichloroethane-d4	5.697	65	49289	5.277	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	105.600%
32) Benzene-d6	5.723	84	191954	5.068	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.400%
36) 1,2-Dichloropropane-d6	6.687	67	58849	5.099	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	102.000%
41) Toluene-d8	7.893	98	177168	4.958	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.200%
43) trans-1,3-Dichloroprop...	8.176	79	20999	4.742	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	94.800%
46) 2-Hexanone-d5	8.629	63	92903	51.646	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	103.300%
56) 1,1,2,2-Tetrachloroeth...	10.751	84	48681	5.068	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	101.400%
66) 1,2-Dichlorobenzene-d4	12.189	152	59672	5.224	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	104.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU032423\
Data File : VU053405.D
Acq On : 24 Mar 2023 21:12
Operator : JC/MD
Sample : 02053-06
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
EYF72

Quant Time: Mar 25 00:42:57 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR032423WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Mar 25 00:39:08 2023
Response via : Initial Calibration

