

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU032423\
 Data File : VU053408.D
 Acq On : 24 Mar 2023 22:24
 Operator : JC/MD
 Sample : 02092-04
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 CB966

Quant Time: Mar 25 00:43:37 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR032423WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Mar 25 00:39:08 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.243	114	124482	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.414	117	119752	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.809	152	58412	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	45330	5.293	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	= 105.800%		
7) Chloroethane-d5	1.912	69	41640	5.358	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	= 107.200%		
11) 1,1-Dichloroethene-d2	2.565	65	20625	5.654	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	= 113.000%		
20) 2-Butanone-d5	4.645	46	108276	55.192	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	= 110.380%		
24) Chloroform-d	5.060	84	85551	5.389	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 107.800%		
26) 1,2-Dichloroethane-d4	5.697	65	44375	5.454	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 109.000%		
32) Benzene-d6	5.723	84	181040	5.335	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 106.600%		
36) 1,2-Dichloropropane-d6	6.684	67	57506	5.561	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	= 111.200%		
41) Toluene-d8	7.893	98	167044	5.217	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 104.400%		
43) trans-1,3-Dichloroprop...	8.176	79	19566	4.930	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	= 98.600%		
46) 2-Hexanone-d5	8.629	63	89307	55.405	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	= 110.800%		
56) 1,1,2,2-Tetrachloroeth...	10.751	84	48101	5.588	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	= 111.800%		
66) 1,2-Dichlorobenzene-d4	12.189	152	59377	5.650	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	= 113.000%		
Target Compounds						Qvalue
51) Chlorobenzene	9.443	112	76911	3.478	ug/L	96
65) 1,4-Dichlorobenzene	11.835	146	12178	0.777	ug/L	94
67) 1,2-Dichlorobenzene	12.208	146	3652	0.247	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU032423\
Data File : VU053408.D
Acq On : 24 Mar 2023 22:24
Operator : JC/MD
Sample : 02092-04
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
CB966

Quant Time: Mar 25 00:43:37 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR032423WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Mar 25 00:39:08 2023
Response via : Initial Calibration

