

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU032423\
 Data File : VU053410.D
 Acq On : 24 Mar 2023 23:13
 Operator : JC/MD
 Sample : 02092-06
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 CB970

Quant Time: Mar 25 00:44:06 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR032423WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Mar 25 00:39:08 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.247	114	146208	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.414	117	135363	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.806	152	66625	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	46164	4.590	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	91.800%	
7) Chloroethane-d5	1.912	69	43168	4.729	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	94.600%	
11) 1,1-Dichloroethene-d2	2.565	65	19475	4.545	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	91.000%	
20) 2-Butanone-d5	4.636	46	118584	51.464	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	102.920%	
24) Chloroform-d	5.057	84	90485	4.853	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	97.000%	
26) 1,2-Dichloroethane-d4	5.697	65	47229	4.943	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	98.800%	
32) Benzene-d6	5.723	84	192807	5.026	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	100.600%	
36) 1,2-Dichloropropane-d6	6.687	67	58601	5.013	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	100.200%	
41) Toluene-d8	7.893	98	171889	4.749	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	95.000%	
43) trans-1,3-Dichloroprop...	8.176	79	19913	4.439	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	88.800%	
46) 2-Hexanone-d5	8.629	63	90430	49.632	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	99.260%	
56) 1,1,2,2-Tetrachloroeth...	10.751	84	47239	4.855	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	97.200%	
66) 1,2-Dichlorobenzene-d4	12.189	152	59307	4.948	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	99.000%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU032423\
Data File : VU053410.D
Acq On : 24 Mar 2023 23:13
Operator : JC/MD
Sample : 02092-06
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
CB970

Quant Time: Mar 25 00:44:06 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR032423WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Mar 25 00:39:08 2023
Response via : Initial Calibration

