

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU032519\
 Data File : VU030417.D
 Acq On : 25 Mar 2019 21:04
 Operator : JC/SP
 Sample : K2016-19
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C09E8

Quant Time: Mar 26 03:02:27 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM032319WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Mar 26 01:25:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	459243	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	434837	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	168360	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	194009	52.49	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	104.98%
7) Chloroethane-d5	1.68	69	150369	51.33	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	102.66%
11) 1,1-Dichloroethene-d2	2.27	63	263324	37.56	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	75.12%
21) 2-Butanone-d5	4.17	46	317177	91.29	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	91.29%
24) Chloroform-d	4.65	84	301868	47.40	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.80%
26) 1,2-Dichloroethane-d4	5.31	65	199697	47.92	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.84%
32) Benzene-d6	5.34	84	621313	49.82	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.64%
36) 1,2-Dichloropropane-d6	6.33	67	218530	51.22	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	102.44%
41) Toluene-d8	7.56	98	539258	48.15	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.30%
43) trans-1,3-Dichloropropene-	7.85	79	91541	45.63	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	91.26%
47) 2-Hexanone-d5	8.31	63	201973	87.36	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	87.36%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	281041	46.03	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	92.06%
64) 1,2-Dichlorobenzene-d4	11.86	152	170210	50.46	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.92%


Target Compounds					Ovalue
13) Acetone	2.32	43	14111	4.016 ug/L	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU032519\
Data File : VU030417.D
Acq On : 25 Mar 2019 21:04
Operator : JC/SP
Sample : K2016-19
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C09E8

Quant Time: Mar 26 03:02:27 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM032319WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Mar 26 01:25:56 2019
Response via : Initial Calibration





#13

Acetone

Concen: 4.016 ug/L

RT: 2.32 min Scan# 383

Delta R.T. -0.00 min

Lab File: VU030417.D

Acq: 25 Mar 2019 21:04

Instrument :
MSVOA_U
ClientSampleId :
C09E8

Tot Ion: 43 Resp: 14111

Ion Ratio Lower Upper

43 100

58 29.8 0.0 62.8

