

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU032521\
 Data File : VU042775.D
 Acq On : 25 Mar 2021 11:00
 Operator : SY/MD
 Sample : VU0325WBL01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VU0325WBL01

Quant Time: Mar 26 02:03:26 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U032421W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 25 05:42:00 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.385	168	112335	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.257	114	171515	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.427	117	165426	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	82262	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.710	65	87844	49.05	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.10%	
35) Dibromofluoromethane	5.298	113	60879	47.56	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.12%	
50) Toluene-d8	7.906	98	215668	47.89	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.78%	
62) 4-Bromofluorobenzene	10.642	95	78497	44.86	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.72%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU032521\
Data File : VU042775.D
Acq On : 25 Mar 2021 11:00
Operator : SY/MD
Sample : VU0325WBL01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU0325WBL01

Quant Time: Mar 26 02:03:26 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U032421W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 25 05:42:00 2021
Response via : Initial Calibration

