

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU032521\
 Data File : VU042778.D
 Acq On : 25 Mar 2021 12:31
 Operator : SY/MD
 Sample : PB1355334TB
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PB1355334TB

Quant Time: Mar 26 02:04:48 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U032421W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 25 05:42:00 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.385	168	115060	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.256	114	182852	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.427	117	182679	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.822	152	85742	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.710	65	90756	49.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.96%	
35) Dibromofluoromethane	5.298	113	63434	46.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.96%	
50) Toluene-d8	7.906	98	229243	47.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.50%	
62) 4-Bromofluorobenzene	10.639	95	85511	45.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.66%	
Target Compounds					Qvalue	
16) Acetone	2.636	43	20182	8.71	ug/l	99
20) Methylene Chloride	3.057	84	6727	5.09	ug/l	95
43) Isopropyl Acetate	5.919	43	14424	3.41	ug/l	99
95) Naphthalene	14.095	128	5010	3.39	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU032521\
Data File : VU042778.D
Acq On : 25 Mar 2021 12:31
Operator : SY/MD
Sample : PB1355334TB
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PB1355334TB

Quant Time: Mar 26 02:04:48 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U032421W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 25 05:42:00 2021
Response via : Initial Calibration

